

# U.S. NUCLEAR REGULATORY COMMISSION TELEPHONE CONVERSATION RECORD

Date:

4/19/2012 / 4:10 PM

Report No. 2011-001

License No.  
37-30850-01Docket No.  
03036438

Name of Licensee: GAMMA IRRADIATOR SERVICE (GIS)

Name of Participant(s): Doyle Terry Stout, President and RSO  
Kathy Modes, NRC

Telephone No. 570-925-5681

Subject: Request letter documenting corrective actions

(NOTE: This will be used as the  
Documents Title in ADAMS)

Licensee will provide documentation of their corrective actions taken and planned as summarized below:

1. On March 8, 2012, GIS ceased all service on Category II irradiators until the license is amended.
2. Informed all GIS personnel about this matter on March 8, 2012
3. On March 9, 2012, trained personnel on the importance of reviewing device documentation to verify type of device to ensure GIS is authorized to perform that service.
4. Plan to submit an amendment request to service JL Shepherd Model 81 Category II irradiators by May 30, 2012, if not sooner.
5. Developed interim guidance on how to respond to emergency events that could be encountered and trained personnel on these interim emergency procedures on April 6, 2012.
6. Plan to submit the revised written emergency procedures as required by the CAL by mid-May 2012.
7. Plan to train personnel on new operating and emergency procedures once NRC approves procedures.

## Action Required: Await for response from licensee

Document Availability:

Publicly Available

Non-Publicly Available

<input checked="" type="checkbox"/>	Non-Sensitive
<input type="checkbox"/>	Sensitive – Privacy Act/ PII
<input type="checkbox"/>	Immediate Release

<input type="checkbox"/>	Non-Sensitive Copyright
<input type="checkbox"/>	Sensitive – Internal
Normal Release Date: 4/27/12	

<input type="checkbox"/>	Sensitive- Proprietary
<input type="checkbox"/>	Sensitive – Security-Related
Delayed Release Date:	

SUNSI Review Completed By: K. Modes

/ RA /

Document Accession #: ML121150447