

**From:** Elliott, Robin  
**Sent:** Tuesday, April 17, 2012 1:45 PM  
**To:** 'Debbief@cardiocppa.com'  
**Cc:** 'jgfong@aol.com'  
**Subject:** Renewal Application for NRC license No. 07-30713-01

Licensee Name: Cardiology Physicians, P.A.  
License No.: 07-30713-01  
Docket No.: 030-35921  
Mail Control No.: 577079

Dear Dr. Pastore:

As per my phone conversation with Jim Fongheiser today on your behalf, additional information is needed to process the renewal application for Cardiology Physicians, P.A. Please provide the following:

- In order to facilitate communications please provide a business email address, if you use one, and fax number.
- Please confirm that you do not possess nor intend to use PET radiopharmaceuticals.
- Provide the manufacturer and model number for any sealed sources that do not meet the criteria in 10 CFR 35.65 (e.g. greater than 30 millicuries).
- The close out survey of previously used areas at the 1400 Foulk Road location was not submitted, as indicated in your letter dated March 24, 2008, after discontinuing use in these areas. Please submit it for our review.
- A closeout survey was not submitted for the camera room, which is now labeled as the echo room, at the Abby Medical Center. Please submit a closeout survey for this location.
- Further additional information is needed for the current facility diagrams submitted as outlined in NUREG 1556 Volume 9 Revision 2 and can be found in section 8.16:  
<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/sr1556v9r2-final.pdf#08-16>

As well as in NUREG Volume 9 Revision 2 Appendix E Figure E.1:  
<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/sr1556v9r2-final.pdf#app-e>

The additional information we are requesting is as follows:

- Show room numbers and doors for the use areas if they exist.
- Show adjacent rooms, what exists above and below the use areas and the relation of the suite to the exterior of the building as applicable.
- Provide information related to the security of the hot lab.
- Drawings and diagrams that provide exact locations of materials or depict specific locations of safety or security equipment should be marked as "Security-related information – withhold under 10 CFR 2.390."

Once we receive the additional information requested, we should be able to finalize the processing of your renewal application. You may respond to my attention in writing by letter or fax (610-337-5269) referencing Mail Control No. 577079. If we do not receive a reply from you within 30 calendar days

from the date of this e-mail, we will assume that you do not wish to pursue your renewal application.  
Please confirm receipt of this communication.

Regards,

Robin L. Elliott  
Health Physicist  
U. S. Nuclear Regulatory Commission  
Region 1  
475 Allendale Road  
King of Prussia, PA 19406  
(610) 337-5076 voice  
(610) 337-5269 fax  
[robin.elliott@nrc.gov](mailto:robin.elliott@nrc.gov)