Application and Review Checklist for (Acceptance, 1st, or 2nd) Review for SSD 12 - 12

SUMMARY DATA							
Name and Complete Mailing Address of the Applicant:		Name, Title, and Telephone Number of the Individual to Be Contacted If Additional Information or Clarification Is Needed by the NRC:					
The Applicant is (check one):			If the Applicant Is Not the Manufacturer, Provide the Name and				
Custom User		Compl	Complete Mailing Address of the Manufacturer:				
	Manufacturer						
	Distributor						
×	Manufacturer and Distributor			4			
If the Applicant Is a Custom User, Provide the Name and Complete Mailing Address of the Distributor:		Provide the Name, Complete Mailing Address, and Function of Other Companies Involved:					
Model	Number:	Principal Use Code (see Appendix C): A= (])					
	Used by the Industry to Identify the	For Use by:					
Device	ct (e.g., Radiography Exposure e, Teletherapy Source, Calibration	Specific Licensees Only					
Source	e, etc.):	General Licensees Only					
		Both Specific and General Licensees					
			Persons Exempt from Licensing				
Leak-T	est Frequency:	Principal Section of the 10 CFR that Applies to the User (e.g.,					
	Periodic Leak-Testing is Not Required	Genera	al Licensees under 10 CFR 31.5):				
X	6 Months	Radionuclides and Maximum Activities (including loading					
	Attached is justification for a leak test frequency of greater than 6 months	tolerance):					
CERTIFICATION:							
THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT. THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30 AND 32 AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF. WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.							
Certifying Officer — Typed Name and Title							
Signature:				Date:			

CHECKLIST

Registration Certificate Holder:

Model:

DESCRIPTION	OK/DEF		COMMENTS
	1 st Reviewer	2 nd Reviewer	
DESCRIPTION/CONSTRUCTION			
If registration certificate holder is requesting to register more than one source/device on a certificate, are designs similar enough to do so?	NP	42	
Device/source design with complete engineering drawings (dimensions, tolerances, list of materials)	det	complete	RAZ
Assembly methods (screw, welds, etc.); verify integrity	0 (/	
Source mounting (size and integrity) and security	46	/	
Is source ANSI classification sufficient (from ANSI N43.6-1997 and ISO 2919): Radiography - Unprotected 43515 Radiography - In Device 43313 Medical - Radiography 32242 Medical - γ Teletherapy 53524 Medical - Brachytherapy 53211 γ Gauges - Unprotected 43333 γ Gauges - Unprotected 43333 γ Gauges - In Device 43232 β Gauges, Low Energy γ Gauges, or X-ray fluorescence 33222 Oil Well Logging 56522 Portable Moist/Density 43333 Neutron Applications 43323 Calibration source activity > 30μCi (1 MBq) 22212 γ Irradiators (II, III) 43424 γ Irradiators (II, III) 43424 Chromatography 32211 Static Eliminators 22222 Smoke Detectors 332222	0		
Definition of shutter operation (locked in "off" position, not locked in "on" position), Fail safe, spacing and tolerances	NA	nla	
On-Off indicators (description, qty., location)	NIA	nla	
Safety interlocks, guards, and so forth to prevent access to beam or high radiation levels	WIA	nh	
Corrosion between unlike materials (e.g., aluminum and steel, depleted uranium & steel, and so forth)	r IA	nla	
Shielding efficiency and integrity	NA	nla	

CHECKLIST

Registration Certificate Holder:

Model:

DESCRIPTION	OK/DEF		COMMENTS
	1 st Reviewer	2 nd Reviewer	
For medical devices:			
Type of FDA approval: Premarket notification (501(k)) Premarket approval (PMA) Investigational Device Exemption (IDE) Humanitarian Device Exemption (HDE)			
Type of Medical Use: manual brachytherapy, 35.400 medical diagnosis, 35.500 photon-emitting remote afterloader, 35.600 photon-emitting teletherapy unit, 35.600 gamma stereotactic radiosurgery unit, 35.600 other medical, 35.1000 (intervascular brachytherapy, beta-emitting remote afterloaders, etc) List of FDA limitations of use provided	NIA	nla	
Well logging (10 CFR 39.41) and irradiator (10 CFR 36.21) sources must be as nondispersible and nonsoluble as practical.	NIA	Mh	P
See "ANSI and Other Standards" list for references for particular source/device designs (e.g. radiography, Brachytherapy, etc.)	of		

LABELING			2
Complete and final copy of label attached	NA	1/4	
Materials, dimensions, colors (note on registration certificate if labeling is exempt from the color requirements of 10 CFR Part 20)	pla	nla	
Attachment and location(s) - visible to users?	NA	nle	
Method of attachment is durable and permanent under normal conditions of use	NA	nh	
Contents: Model#, Serial#, Isotope, Activity, Manufacturer, Date of Assay, Trefoil, "CAUTION – RADIOACTIVE MATERIAL" (Depleted Uranium information must be included)	. rla	~14	nodayl
Is label in compliance with regulatory requirements?			
CONDITIONS OF USE			
Estimated working life of the source/device (years, operational cycles)	qula		ho cheryl
Actions to be taken when product reaches end of its working life.	,	Ma	
Maximum allowable temperature, vibration, shock, corrosion, etc. (during use, handling, storage, and transport)		nla	,
How the device will be used		nla	3
Meets dose limits of Part 32 for distribution general licensees or persons exempt from licensing		nla	
PROTOTYPE TESTING/HISTORICAL USE			
Tests methods and conditions (for source and device)	6 ¢		
Tests results	oK		
Years of use (incidents, failures, etc.)	NIA	2/5	
Similarities to other sources/devices if they are used as basis.	i /A	nla	
RADIATION PROFILES			
Survey instrument used (type, window thickness, sensitivity, calibration dates, etc.)	ρK	/	
Conditions: including environments, scatter (product in beam), and use of guards and shields	, /A	nla	
RADIATION PROFILES (CONTINUED)			
Distance from source/surface (per ANSI 538-1979, N43.8 - 2001)	υK	/	
Shutter Open and Closed/Source Shielded	n ha	nla	
Verify radiation surveys for γ radiation meet inv² law.	/	/	

Verify radiation surveys for non-γ radiation have not been calculated using inv ² law.	e/		
QUALITY ASSURANCE			
Materials, subassemblies, services		✓	
Assembly methods (screws, welding, etc.)		nla	
Dimensions and tolerances		nla	Same
Activity, radiation levels, leak tests	NCH	nla	>
Final inspection		n/L	
QA Manual and comparison of other (generally) accepted guidance (e.g., ANSI/ISO/ASQ 9001-2001)		nla	
Additional measures for SSDs if ANSI/ISO/ASQ 9001-2001 is used		n/L	
INSTALLATION			
Fixed, portable, movable, fixed installation but portable source housing	7 110-	Ma	
Inherent shielding, inaccessibility	(0)	7/6	,
Beam access: size of air gap/opening to beam and use of interlocks, locks, additional shielding or barriers	/	nle	
Mounting integrity		nla	
ACCOMPANYING DOCUMENTATION			9
Leak tests results and radiation surveys	^	nle	
Operation safety instructions, maintenance, calibration, damage/failure, specific warnings, leak test, and radiation survey instructions if applicable	NA	nla	
For Distribution to General Licensees: Verify NRC Regions and Agreement State listing is up- to-date and copies of all pertinent regulations		nla	

he following activities may be performed by the ersons indicated: Activity by a Only by a Will be General Specific Offered		v.	
General Specific Offered			
Licensee Licensee by the Applicant			
nstallation			
relocation			
laintenance			Some
repair	TIA		
ource xchange			
alibration			
eak Testing		. "	
dadiation durvey			**
raining			
FOREIGN VENDORS	0		
Prop ship	NA		0
Who and where is source installed	ALL AND PROPERTY AND PARTY.		
eak test and radiation surveys			
A in the U.S.	U .		

1 st Reviewer Signature: _	J-5,10	Date:	
2 nd Reviewer Signature:	Skyngh	<u>Date:</u>	Ayor 200 2012