

**U.S. NUCLEAR REGULATORY COMMISSION
TELEPHONE CONVERSATION RECORD**

Date: 5/17/2012

Time:

Mail Control or Report No(s). 577245 License No(s). 44-10187-03 Docket No(s). 03003289

Name of Licensee: Fletcher Allen Health Care

Name of Participant(s): Marleen Moore, RSO/AMP

Telephone No. 802-847-3506, e-mail Marleen.Moore@vtmednet.org

Subject: Additional information for amendment request

Fletcher Allen's request to upgrade the high dose rate remote afterloader (HDR) to a new model included updated emergency procedures. The procedures do not appear to address all of the information described in 10 CFR 35.610(a)(4). Fletcher Allen's former HDR emergency procedures submitted Feb. 13, 2003, included on page 23 a sheet called "Gammamed 12i Emergency Procedure" to be posted on the wall next to the console. It briefly summarized the position titles and duties of the individuals responsible to initiate emergency actions, the process for restricting access/posting the treatment area, and names/telephone numbers of RSO/AMP/AU to be contacted in an emergency. I asked Marleen if Fletcher Allen plans to continue to use this procedure for the new GammaMed.

Marleen said that Fletcher Allen plans to continue to use this procedure, with minor modifications. She agreed to submit a copy to be considered as part of the amendment request.

Marleen also asked for the amendment to be issued on an expedited basis. Varian wishes to upgrade the HDR at the next source replacement scheduled for 6/20/2012, and requested a copy of the amended license 30 days in advance.

Action Required: Await response, then issue amendment on expedited basis

Document Availability:

☒

Publicly Available

☐

Non-Publicly Available

☒

Non-Sensitive

☐

Sensitive – Privacy Act/ PII

☐

Immediate Release

☐

Non-Sensitive Copyright

☐

Sensitive – Internal

☐

Normal Release Date:

☐

Sensitive- Proprietary

☐

Sensitive – Security-Related

☐

Delayed Release Date:

SUNSI Review Completed By: SGabriel / RA /

Document Accession #: ML #12139A162

