

Sara A.B. Forster  
MATERIALS LICENSING BRANCH

TELECON & FAX TRANSMITTAL



TO: file  
COMPANY: N/A

NUCLEAR REGULATORY COMMISSION  
REGION III  
2443 WARRENVILLE ROAD  
LISLE, ILLINOIS 60532-4351

# PAGES: 2 TEL.: N/A

FAX: N/A  
EMAIL: N/A

(630) 829-9892 FAX: (630) 515-1078

---

**CONVERSATION RECORD**

|TIME |DATE  
1:30 pm April 24, 2012

---

NAME OF PERSON(S) CONTACTED	TELEPHONE NO.	ORGANIZATION
Arthur Jerome Ewald, M.S., AMP	(810) 342-3805	McLaren – Flint

---

REPRESENTED PERSON or PERSONS	ORGANIZATION
Christopher J. Conlin, MD, Radiation Safety Officer	McLaren – Flint

---

SUBJECT
License No.: 21-04171-04  Control No.: 577008

---

**SUMMARY**

We have reviewed your license amendment request and find that we are unable to continue this action until we have received information regarding the following:

1. The letter dated 2/13/2012 indicated that your current HDR would be replaced by a newer model. Please note that both the current and new HDR models will be listed on your license until the old model has been removed from your facility. Once the old HDR model is removed, and the NRC receives documentation confirming the same, the old model number will be removed from the license. Documentation should include final leak test results, date of transfer, and confirmation by the manufacturer that the older model sealed sources were received, if applicable.

**RESPONSE:** In the phone conversation, the licensee contact indicated that the HDR model number change reflects changes to the control room console only; the sources and device will not change. The licensee will submit confirmation of the console change upon its completion. No additional information is needed at this time.

2. The revised HDR Emergency procedures were cumbersome and do not appear to be immediately useful in the event of an actual emergency. Please resubmit an emergency procedure that would be posted on the HDR console, and would actually be used in an emergency. Refer to NUREG 1556, Vol. 9, Rev. 2, page 8-51, for additional guidance on this request.

**RESPONSE:** An updated HDR emergency procedure was submitted via facsimile dated 04/24/2012. No additional information is required.

---

We have requested that you submit the referenced item:

- HDR emergency procedures (information provided was sufficient.)

- via facsimile, to (630) 515-1078. Please reference the Control No. 577008, as listed at the top of this memo. We expected to, and did, hear from you before May 8, 2012. No additional information is required.

**For future reference, please always include the name, phone number and fax number of at least one person whom we may contact for additional information when reviewing your licensing correspondence and requests.**

Please submit the requested information within 14 days of this record. **Include reference control number 577008, Please FAX your response to my attention at (630) 515-1078.** You may also scan your response and send to me via email, as a pdf file.

Please direct any questions you have to me at **(630) 829-9892** or **sara.forster@nrc.gov**.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Sara A.B. Forster

*Sara A.B. Forster* 05/10/2012  
#