



West Virginia University

Radiation Safety Department

May 29, 2012

Docket No. 030-20233

License No. 47-23066-02

Control No. 576453

Tara Weidner, Health Physicist
United States Nuclear Regulatory Commission-Region I
475 Allendale Road
King of Prussia, Pennsylvania 19406

SUBJECT: Response to U.S. NRC E-mail request, on May 25, for additional information

Dear Ms. Weidner:

This is in reference to your e-mail request for additional information dated May 25, 2012.

1. We request that item 6.A. be written as follows:
Any byproduct material with atomic numbers 3 through 83 and half life of less than 120 days
1 curie per radionuclide and 15 curies total
2. Survey instrument utilized at West Virginia university Hospitals will be calibrated at prescribed frequency with record of verification that indicates the procedure used and the data obtained. And also it includes the owner or user of the instrument, a description of the instrument, including the manufacturer's name, model number, and serial number, description of the calibration source, including the exposure rate at a specified distance or activity on a specified date, and also the name of the person who performed the calibration and the date it was performed. A listing of all instruments is maintained at the Radiation Safety Office's Central Database System. The database is checked periodically by Radiation Safety Office for calibration due dates to assure that all survey instruments are not out -of-calibrations. The calibration of all instrument and procedures will be reviewed by the RSO during subsequent inspections and audits and it will be reported during the RSC quarterly meetings.

**Robert C. Byrd Health Sciences Center
West Virginia University
WVU Hospitals**

G-139 Health Sciences North
PO Box 9006
Morgantown, WV 26506-9006

Phone: 304-293-3413
Fax: 304-293-4529

Equal Opportunity/Affirmative Action Institution



West Virginia University

Radiation Safety Department

If you have any further questions or concerns regarding our responses please feel free to contact the Radiation Safety Department at 304-293-1554.

Sincerely,

A handwritten signature in black ink, appearing to read 'Nasser Razmianfar', written over a horizontal line.

Nasser Razmianfar
Director and Radiation Safety Officer

**Robert C. Byrd Health Sciences Center
West Virginia University
WVU Hospitals**

G-139 Health Sciences North
PO Box 9006
Morgantown, WV 26506-9006

Phone: 304-293-3413
Fax: 304-293-4529

Equal Opportunity/Affirmative Action Institution