

**From:** [Gabriel, Sandra](#)  
**To:** ["Price, Kenneth W."](#)  
**Cc:** ["Fomenko, James"](#)  
**Subject:** Additional information for NRC license amendment request, mail control 577516  
**Date:** Tuesday, June 05, 2012 8:34:00 PM

---

Licensee: University of Connecticut Health Center  
License Number: 06-13022-02  
Docket Number: 03001295  
Mail Control: 577516

To: Kenneth W. Price, M.P.H., C.H.P., Radiation Safety Officer

Additional information is needed for the NRC to continue its review of your amendment request to remove Building 27 from your license. Please provide the following information within 30 days:

- 1) Confirm that all equipment was removed from Building 27 and all licensed material was moved to other locations on campus or transferred to an authorized recipient. For any transferred sealed sources, provide documentation that they were not leaking and were received by the authorized recipient.
- 2) The description of operational history of Building 27 mentioned U-238 as uranyl nitrate and uranyl acetate. Please explain why these were not included in the list of possible contaminants. For example, were these dry solids in their original bottles, and the bottles were never opened in Building 27?
- 3) Your request stated that no wipe test was observed to have a detectable amount of radioactivity present and all meter surveys indicated no reading above the detection limit, however the individual results were not submitted. Provide all actual data from the final status surveys, including identification of instruments used and calibration certificates.

Please provide a written response to these items within 30 days. You may provide this to my attention at the new Region I office, referencing mail control 577516.

Please contact me by e-mail with any questions. I am currently on an inspection trip until June 11 and will check e-mail each evening.

Thank you,  
Sandy Gabriel, Ph.D.  
Senior Health Physicist  
Medical Branch  
NRC Region I  
610-337-5182 (voice)  
610-337-5269 (fax)