

U.S. NUCLEAR REGULATORY COMMISSION
TELEPHONE CONVERSATION RECORD

Date: 6/12/2012

Mail Control 577332 **License No(s).** 41-06832-06 **Docket No(s).** 03017797
or Report No(s).

Name of Licensee: Tennessee Valley Authority
Name of Participant(s): Lee Miller, TVA
Randolph Ragland, NRC RI,
Telephone No. 423-751-3197; 610-337-5083
Subject: RAI License Renewal Application
(NOTE: This will be used as the Documents Title in ADAMS)

I noted that the cover letter dated March 22, 2012, for the license renewal application included a statement that "There are no regulatory commitments made by this submittal," was contrary to the license renewal process, in that NRC specifically requests licensee commitments as part of the license renewal process (e.g., the application gets tied down as a license condition). In addition, I noted that the application did not include a source misconnect test in the quarterly maintenance process and alerted Mr. Miller to the Sentinel Service Bulletin 2004 which recommends this test be added to the quarterly maintenance procedures..

Action Required: TVA will send a letter which indicates that TVA recognizes that the March 22, 2012, license application represents a regulatory commitment.

Document Availability: **Publicly Available** **Non-Publicly Available**
 Non-Sensitive **Non-Sensitive Copyright**
 Sensitive – Privacy Act/ PII **Sensitive – Internal** **Sensitive- Proprietary**
 Immediate Release **Normal Release Date: 6/20/2012** **Sensitive – Security-Related**
Delayed Release Date:
SUNSI Review Completed By: Randolph C. Ragland, Jr. / RA /