

# Cline's Technical Services, Inc.

10883 Cincinnati Zanesville Road - Amanda, Ohio 43102 - Ph 740.969.2720 Fax 740.969.2122

## Leak Test Report

S & M E, Inc.  
6190 Enterprise Ct.  
Dublin, OH 43016

Model Number: TRX 3401

Serial Number: 6971

Isotope: Cs-137 8 mCi  
Am-241/Be 40 mCi

Date of test: 05-21-2012

The sample and information above was submitted for leak test analysis.

Analysis date: 05-23-2012

Analysis Number: 1774

Results: Removable Contamination

< 7 Bq

Becquerel "Bq"

Note:

The US Nuclear Regulatory Commission and Agreement States require that the analysis of the wipe from a sealed source be capable of detecting the presence of 185 Becquerel ( or 0.005 micro curie ) on the sample. { The result of this wipe is below 185 Bq. } The Source is not considered leaking. This report should be retained for viewing by regulatory agencies.

( < = less than )

Signed Willie Cline  
Office of Radiation Safety  
Ohio Materials License No. 03225230000