

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

*0202500 577681*

<p>Licensee</p> <p>1. PETNET Solutions, Inc.</p> <p>2. 810 Innovation Drive Knoxville, TN 37932</p>	<p>In accordance with letter dated <b>June 8, 2012,</b></p> <hr/> <p>3. License number 41-32720-02MD is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration date <b>September 30, 2020</b></p> <hr/> <p>5. Docket No. <b>030-38158</b> Reference No.</p>
---	---

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Fluorine-18	A. Any	A. 20 curies
B. Nitrogen-13	B. Any	B. 2 curies
C. Carbon-11	C. Any	C. 4 curies
D. Oxygen-15	D. Any	D. 2 curies
E. Hydrogen-3	E. Liquid	E. 5 millicuries
F. Any byproduct material authorized under 10 CFR 35.65	F. Sealed sources	F. 200 millicuries

9. Authorized use:
- A. through D. Preparation and distribution of radioactive drugs in accordance with 10 CFR 32.72. Preparation and distribution of radioactive drugs and radiochemicals to authorized recipients for nonmedical use.
  - E. Possession incidental to the production of radionuclides.
  - F. Calibration and checking of the licensee's instruments.

CONDITIONS

10. Licensed material shall be used only at the licensee's facilities located at 1345 West 16<sup>th</sup> Street, Indianapolis, Indiana.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
41-32720-02MD  
Docket or Reference Number  
030-38158

Amendment No. 02

11. Licensed material shall be used by, or under the supervision of:
- A. A pharmacist working or designated as an authorized nuclear pharmacist in accordance with 10 CFR 32.72(b)(2)(i) and (4).
  - B. Authorized Nuclear Pharmacist(s): James Kauchak, R.Ph., Jared Driskill, R. Ph., Victor Calonico, R.Ph., Ryan Kuenkel, R.Ph, Rick Grigson, R.Ph. and Virginia Coakley, R.Ph.
12. The Radiation Safety Officer for this license is Rick Grigson, R.Ph.
13. This license does not authorized distribution to persons exempt from licensing.
14. A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by NRC under 10 CFR 32.210 or by an Agreement State.
- B. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by NRC under 10 CFR 32.210 or by an Agreement State prior to the transfer, a sealed source or detector cell received from another person shall not be put into use until tested.
- C. Sealed sources need not be tested if they are in storage and are not being used. However, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- D. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 Becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 Becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.
15. Sealed sources containing licensed material shall not be opened or sources removed from source holders by the licensee, except as specifically authorized.
16. The licensee shall conduct a physical inventory every six months, or at other intervals approved by NRC, to account for all sealed sources and/or devices received and possessed under the license.
17. Except for maintaining labeling as required by 10 CFR Part 20 or 71, the licensee shall obtain authorization from NRC before making any changes in the sealed source, device, or source-device combination that would alter the description or specifications as indicated in the respective Registration Certificates issued either by the Commission pursuant to 10 CFR 32.210 or by an Agreement State.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number

41-32720-02MD

Docket or Reference Number

030-38158


Amendment No. 02

18. The licensee is authorized to hold byproduct material with a physical half-life of less than or equal to 120 days from decay-in-storage before disposal without regard to its radioactivity if the licensee:
- A. Monitors byproduct material at the surface before disposal and determines that its radioactivity cannot be distinguished from the background radiation level with an appropriate radiation detection survey meter set on its most sensitive scale and with no interposed shielding;
  - B. Removes or obliterates all radiation labels, except for radiation labels on materials that are within containers and that will be managed as biomedical waste after they have been released from the licensee;
  - C. Maintains records of the disposal of licensed materials for 3 years. The record must include the date of the disposal, the survey instrument used, the background radiation level, the radiation level measured at the surface of each waste container, and the name of the individual who performed the disposal.
19. The licensee is authorized to retrieve, receive, and dispose of radioactive waste from its customers limited to radiopharmacy-supplied syringes and vials and their contents.
20. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
21. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- B. Application dated September 4, 2009; and
  - C. Letters dated September 11, 2009, February 17, 2010, May 24, 2011, July 14, 2011, and **June 8, 2012.**

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date JUN 14 2012

By

  
Kevin G. Null  
Materials Licensing Branch  
Region III