



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
1600 EAST LAMAR BLVD
ARLINGTON, TEXAS 76011-4511

July 6, 2012

MEMORANDUM TO: Angela McIntosh, Health Physicist (T-8E24)
Radioactive Materials Safety Branch
Division of Materials Safety & State Agreements

FROM: G. Michael Vasquez, Chief */RA ADG for/*
Nuclear Materials Safety Branch A
Division of Nuclear Materials Safety

SUBJECT: MATERIALS LICENSEE EVENT REPORT SUMMARY

Enclosed is the subject report for June, 2012. This enclosure documents the licensee event reports that were closed the preceding month, and lists ADAMS ML numbers for documents associated with these events, detailing Region IV's response to these events.

Enclosure:

June, 2012 RIV DNMS Licensee
Event Reports Summary Table

cc via email w/Enclosure 1:
E. Collins
A. Howell
A. Vegel
V. Campbell
J. Whitten
D. Spitzberg
J. Thompson
D. Huntsman [Dante.Huntsman@inl.gov]
LER File

S:\DNMS\INMIB\LERs\2011\6-2012.doc

ADAMS	<input type="checkbox"/> No	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> SUNSI Review Complete	Reviewer Initials: JLT
	<input checked="" type="checkbox"/> Publicly Available		<input checked="" type="checkbox"/> Non-sensitive	
	<input type="checkbox"/> Non-publicly Available		<input type="checkbox"/> Sensitive	
KEYWORD: SUNSI Review Complete				
RIV:DNMS/NMSB-A	C:NMSB-A			
JLThompson	GMVasquez			
/RA ADG for/	/RA ADG for/			
07/05/2012	07/05/2012			

OFFICIAL RECORD COPY

T=Telephone

E=E-mail

F=Fax

RIV DNMS LICENSEE EVENT REPORTS SUMMARY TABLE - June 2012								
ADAMS ML # or Event #	Event Date	Notification Date	Licensee	License Number	Reporting Requirement	Event Description	Assigned Inspector	Review Status
47309	9/30/2011	10/3/2011	United States Air Force	42-23539-01AF	10 CFR 20.2201	Lost americium-241 sealed source (12 millicurie)	Cook	Open
47579	1/5/2012	1/9/2012	Benefis Health Care	25-12710-01	10 CFR 35.3045	90% patient underexposure during esophageal treatment with iridium-192 HDR unit	Gaines	Open
47996	6/5/2012	6/5/2012	Inberg-Miller Engineers	49-19477-01	10 CFR 30.50	Troxler portable nuclear gauge was damaged at a temporary jobsite when run over in Cheyenne, Wyoming	Munoz	Open
47932 ML12185A226	5/16/2012	5/17/2012	The Boeing Company	General Licensee	10 CFR 20.2201	Boeing lost 4 polonium-210 static eliminator devices in Haleakala, Hawaii	Stearns	Opened & Closed
ML12136A139 ML12185A082	4/24/2012	5/7/2012	National Oceanic & Atmospheric Administration	46-23463-01	10 CFR 30.50	Licensee reported a leaking nickel-63 source in a gas chromatograph	Hanson	Opened & Closed