


Radiation Protection Services, Ltd.
Serving Radiology, Nuclear Medicine, Therapy and Industry



1405 Stevenson Dr., Suite 11
Springfield, Illinois 62703

Phone: (217) 786-3839
Fax: (217) 786-3110

July 6, 2012

Control Number: 578780

RAM. license # 24-32384-01

U.S. Nuclear Regulatory Commission
Region III
Materials Licensing Section
2443 Warrenville Road
Suite 210
Lisle, IL 60532

Dennis O'Dowd:

With regard to our phone conversation this morning related to the RAM license renewal application for St. Heart and Vascular, the following is submitted for your review:

Under item 9 on page 6 of the application, the licensee requests to reserve the right to upgrade survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used

Under item 9, Dose calibrator, the licensee assures that equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions.


If you require further information or if I can provide other assistance, please contact my office.

Thank you.


Sincerely,

Thomas Schlarman, Radiation Physicist
Radiation Protection Services, Ltd.
(618)210-3925

cc: Candace Candela, C.N.M.T.
St. Louis Heart and Vascular
11155 Dunn Road - Ste 304 E
St. Louis, MO 63136



Radiation Protection Services, Ltd.
Serving Radiology, Nuclear Medicine, Therapy and Industry



1405 Stevenson Dr., Suite 11
Springfield, Illinois 62703

Phone: (217) 786-3839
Fax: (217) 786-3110

FAX

To: <i>Dennis O'Dowd</i>	From: Tom Schlarman
Fax: <i>630-515-1070</i>	Pages: <i>2</i> (including cover)
Phone:	Date: July 6, 2012
Re: <i>RAM Control # 576780</i>	cc:

Comments:

Please see the attached letter.

Please call me if you have questions or require further information.
Thank you.

Tom Schlarman
Phone: (618) 210-3925
Fax: (618) 473-3528

This transmittal is intended only for the above named recipient. If this fax is received in error, please destroy it. For problems with this fax, please contact phone number 618-210-3925.