



July 12, 2012

Dennis P. O'Dowd
Health Physicist
UNITES STATES NUCLEAR REGULATORY COMMISSION
Region III, Materials Licensing Section
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

Re: Clarification for control number: 577513

Mr. O'Dowd,

Please replace the document dated 5/9/2006 with the faxed document date 03/29/2012. The original document was sent in error. The faxed document is the individual results from the wipe tests conducted on 03/29/2012. The data corresponds to the map that was included in the original request.

Please note the sealed source inventory and leak tests were included as informational documentation only. There were no sealed sources stored in the treadmill area.

Thank you for your cooperation with this matter. If you have any question please contact me at (734) 662-3197.

Sincerely,

A handwritten signature in black ink that reads "Michelle L. Kritzman". The signature is written in a cursive style.

Michelle L. Kritzman, M.S.
Radiological Physicist
ABR Certified, Medical Nuclear Physics



Community Health Center of Branch County
License No. 21-16642-01

Date: 3/29/2012 **Performed By:** Linda Brown, RSO
Survey Meter: Ludlum 14C **Check Value:** 6.5 mR/hr

		cpm	net dpm	mR/h
1	Desk	653	0	0.02
2	Floor	703	100	0.02
3	Floor	703	100	0.02
4	Floor	703	100	0.02
5	Floor	703	100	0.02
6	Floor	703	100	0.02
7	Floor	703	100	0.02
8	Table	668	13	0.02
	Background	663		0.02

ACTION LEVELS: 2000 dpm Conversion factor 2.5 dpm/cpm

COMMENTS: No evidence of removable contamination.



TELECOPY TRANSMITTAL

DATE: 07/13/12

TO: Dennis O'Dowd

Fax Number: 630-515-1078

FROM: Michelle Kritzman

RE: Additional Information for control #577513

COMMENTS: _____

This transmittal consists of 3 pages (including transmittal cover sheet). Please contact me if you have any questions or problems with this transmittal.