

**U.S. NUCLEAR REGULATORY COMMISSION
TELEPHONE CONVERSATION RECORD**

Date: 7/19/2012

Time:

Mail Control
or Report No(s).

577757

License No(s).

44-10187-03

Docket No(s).

03003289

Name of Licensee: Fletcher Allen Health Care
Name of Participant(s): Marleen Moore, RSO/AMP
Telephone No. 802-847-3506, e-mail Marleen.Moore@vtmednet.org
Subject: Request for additional information

The request to add Dr. Walsh as authorized user included some inconsistencies in the Form 313A(AUT). Please resolve the following and resubmit an updated copy signed by a qualified authorized user:

- a. In the "Requested Authorization" section at the top of the first page, the only box checked was the one for I-131 greater than 33 mCi. It appears that you also want Dr. Walsh to be authorized for smaller quantities of I-131. If so, please also check the first box after the "OR."
- b. In the preceptor attestation, "First Section," the box checked was "Board Certification" in the section headed "For 35.390." ABR in diagnostic radiology is approved for 35.392 and 35.394, but not for 35.390. Please complete the 35.392 and 35.394 sections instead of the 35.390 section.
- c. In the preceptor attestation, "Second Section," the only box checked was for I-131 greater than 33 mCi. In the "Third Section," the only box checked was for I-131 less than or equal to 33 mCi. Please check both boxes in both sections.

Action Required: Await response, then issue amendment

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive
 Sensitive – Privacy Act/ PII
 Immediate Release

Non-Sensitive Copyright
 Sensitive – Internal
Normal Release Date:

Sensitive- Proprietary
 Sensitive – Security-Related
Delayed Release Date:

SUNSI Review Completed By: SGabriel / RA /

Document Accession #: _____