

Villar, Sheryl

From: Elliott, Robin
Sent: Thursday, July 19, 2012 3:43 PM
To: cardiovascularspecialists@hotmail.com
Cc: wmlairmore@optonline.net
Subject: License renewal 08-19630-01

Licensee: Anthony E. Watkins, M.D.
License Number: 08-19630-01
Docket Number: 030-19011
Mail Control Number: 577627

Dear Dr. Watkins,

As per my phone conversation with Michael Lairmore on your behalf today, additional information is needed to process the license application for Anthony E. Watkins, M.D.

The additional information we are requesting is as follows:

- To assist in future communications, please provide your business fax number and email address, if they are available.
- Please confirm that you do not possess nor intend to use PET radiopharmaceuticals.
- Provide the manufacturer and model number for any sealed sources that do not meet the criteria in 10 CFR 35.65 (e.g. greater than 30 millicuries).
- Please clarify that radiation monitoring devices will be calibrated by a person and/or company that is qualified to perform survey meter calibrations.
- Please confirm that you will implement and maintain written procedures for the safe use of unsealed radioactive material that meets the requirements of 10 CFR 20.1101 and 10 CFR 20.1301.
- Further additional information is needed for the current facility diagram submitted as outlined in NUREG 1556 Volume 9 Revision 2 and can be found in section 8.16:

<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/sr1556v9r2-final.pdf#08-16>

As well as in NUREG Volume 9 Revision 2 Appendix E Figure E.1:

<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/sr1556v9r2-final.pdf#app-e>

The additional information we are requesting is as follows:

- Show room numbers and doors for the use areas if they exist.
- Show adjacent rooms, what exists above and below the use areas and the relation of the suite to the exterior of the building as applicable.
- Provide information related to the security of the hot lab.
- Drawings and diagrams that provide exact locations of materials or depict specific locations of safety or security equipment should be marked as "Security-related information – withhold under 10 CFR 2.390."
- Please note that detailed procedures were not required to be submitted in support of your application (e.g., dose calibrator calibration procedures) and were not reviewed. These procedures will be reviewed during future inspections

Once we receive the additional information requested, we should be able to finalize the processing of your license application. You may respond to my attention in writing by letter or fax (610-337-5269), referencing mail control 577627. If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your amendment.

Please note that the office of the Region I USNRC Division of Nuclear Materials Safety has moved effective May 9, 2012. Our new address is:

U. S. Nuclear Regulatory Commission
Region I
2100 Renaissance Blvd, Suite 100
King of Prussia, PA 19406-2713

Confirmation of receipt of this email is requested.

Regards,

Robin L. Elliott

Health Physicist
U.S. NRC Region I
2100 Renaissance Blvd.
King of Prussia, PA 19406
(610) 337-5076 voice
(610) 337-5269 fax
robin.elliott@nrc.gov



Think Green - Not every email needs to be printed.