


Synergy Mobile Imaging, Inc.
License No. 21-32673-01
Attn: Maria Rabara or Terry Leewald


I have received your close-out survey sent by facsimile dated May 22, 2012, to terminate your NRC license; however, I need additional information before I can terminate your license. Please respond to the items stated below:

1. Please identify the instruments used to perform the close-out survey and their calibration dates.
2. If direct radiation measurements were performed of the nuclear medicine facility, please provide the results of those surveys.
3. Please provide the background measurements for the swipes and for the direct radiation measurements (surveys).
4. Please describe exactly what radioactive material was sent to Radiological Physics Service, Inc. Please include the last leak test results.

Please respond in writing to each of the above items via email to by June 18, 2012. You may scan your response with a signed and dated cover letter and email to toye.simmons@nrc.gov or you may fax your response to 630-515-1078. Please reference the control number 577526 in your response. If you have any questions, please contact me at 630-829-9842.



Toye L. Simmons



Date

Simmons, Toye

From: Simmons, Toye
Sent: Monday, June 04, 2012 2:37 PM
To: 'synergymobileimaging@gmail.com'
Subject: REQUEST FOR TERMINATION OF NRC LICENSE 21-32673-01
Attachments: Scan001.PDF

In order to terminate Synergy's NRC license additional information is required. Please see the attached.