

Roldan, Lizette

From: Roldan, Lizette
Sent: Tuesday, July 03, 2012 10:14 AM
To: 'todd.christensen@crmcwy.org'
Subject: REQUEST FOR ADDITIONAL INFORMATION REGARDING AMENDMENT CONTROL 577572

License No.: 49-01380-01
Docket No.: 030-03496
Control No.: 577572

Dear Mr. Christensen:

This is in reference to your letter dated May 11, 2012 to amend Nuclear Regulatory Commission License No. 49-01380-01. In order to continue our review, we need the following additional information:

1. Please confirm the shielding recommended in the Memo dated October 5, 2011 – Revised 10/12/11 has been installed in the diagram provided for the new location use.
2. Please confirm you will be meeting the member of the public dose limit for unrestricted areas.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Regulations, Guidance, and Communications**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 7:00 a.m. to 6:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region IV Office and refer to Mail Control No. 577572. You may respond by regular mail, fax at 817-200-1188 or email as a PDF attachment. If you have any technical questions regarding this deficiency letter, please call me at 817-200-1596.

If we do not receive a reply from you by July 31, 2012 we will assume that you do not wish to pursue your application and this amendment request will be voided without prejudice to the resubmission to a new amendment request.

Sincerely,

Lizette Roldán-Otero, Ph.D.

Health Physicist
U.S. Nuclear Regulatory Commission
1600 E. Lamar Blvd.
Arlington, TX 76011-4511
Office: 817-200-1596
Fax: 817-200-1188



Cheyenne Regional
Medical Center

214 EAST 23RD STREET
CHEYENNE, WY 82001
307-634-2273
WWW.CRMWCWY.ORG

RECEIVED

JUL 03 2012

DNMS

License No.: 49-01380-01
Docket No.: 030-03496
Control No.: 577572

Dr. Roldan-Otero

I am writing to inform address the questions received from you by e-mail.

1. Please confirm the shielding recommended in the Memo dated October 5, 2011 – Revised 10/12/11 has been installed in the diagram provided for the new location use.

We have installed the recommended shielding in this described location according to the shielding diagram. This shielding was inspected by myself during the construction process and will be verified/measured to insure public dose safety levels. This verification will be recorded and kept on record at our facility.

2. Please confirm you will be meeting the member of the public dose limit for unrestricted areas.

Yes, as a condition of our license we have agreed to meet the public dose limits for unrestricted areas. These will also be met for the PET/CT area of use.

I hope this sufficiently answers your questions. If you would need additional information please feel free to contact me at (307) 633-7838 or by e-mail at todd.christensen@crmcwy.org.

Todd A. Christensen MS DABR
Authorized Medical Physicist
Radiation Safety Officer
Cheyenne Regional Medical Center
Department of Physics
214 East 23rd Street
Cheyenne WY 82001
Phone: (307) 633 7838
Fax : (307) 633 3716
Todd.christensen@crmcwy.org

Todd A. Christensen 7-3-12