

From: [Elliott, Robin](mailto:Robin.Elliott@heartandvascularclinic.com)
To: aparikh@heartandvascularclinic.com
Cc: ahenry@keystonephysics.com
Subject: Request to add additional location of use
Date: Thursday, August 02, 2012 9:36:00 AM

Licensee: Heart & Vascular Clinic
License Number: 07-31394-01
Docket Number: 030-38183
Mail Control: 577943

Dear Dr. Parikh,

As per my phone conversation yesterday with your consultant, Adam Henry, additional information is needed to process your amendment request.

- Please clarify whether to wish to restrict license activities at the new location to only 35.200. If you retain the flexibility to use 35.100 materials, you would not need to request a license amendment, should you decide to use them at this location in the future.
- Please provide the model and serial number of your Cesium 137 vial reference source if the activity exceeds the limits allowed by 10 CFR 35.65 (30 mCi).
- Please verify the survey meter you have purchased is consistent with the commitment made regarding radiation monitoring equipment in your license application submitted on November 12, 2009.
- Further additional information is needed for the current facility diagram submitted as outlined in NUREG 1556 Volume 9 Revision 2 and can be found in section 8.16: <http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/sr1556v9r2-final.pdf#08-16> or NUREG Volume 9 Revision 2 Appendix E Figure E.1: <http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/sr1556v9r2-final.pdf#app-e>

The additional information we are requesting is as follows:

- Show room numbers and doors for the use areas if they exist.
- State what exists above and below the use areas and show the relation of the suite to the exterior of the building as applicable.
- Provide information related to the security of the hot lab area. In specific, please address how all access locations to the hot lab are secured.
- Drawings and diagrams that provide exact locations of materials or depict specific locations of safety or security equipment should be marked as "Security-related information – withhold under 10 CFR 2.390."

Once we receive the information requested, we should be able to finalize the processing of your license amendment. You may respond to my attention in writing by letter or fax (610-337-5269), referencing mail control 577943. If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your license action.

Please provide confirmation of receipt of this email.

Regards,

Robin L. Elliott

Health Physicist

U.S. NRC Region I

2100 Renaissance Blvd.

King of Prussia, PA 19406

(610) 337-5076 voice

(610) 337-5269 fax

robin.elliott@nrc.gov



Think Green - Not every email needs to be printed.