

UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

JUL 25 2012



Robert A. DeWeese, M.D.
Radiation Safety Officer
Columbus Imaging Associates
d/b/a Columbus Diagnostic Imaging
790 Creekview Drive
Columbus, IN 47201

Dear Dr. DeWeese:

Enclosed is Amendment No. 08 renewing your NRC Material License No. 13-26797-01 in accordance with your request.

This also refers to the telephone conversation record transmitted from me to Rhonda Lennon on June 26, 2012; and to the telephone discussions and emails between Ms. Lennon and/or Will Breeden and me on various dates, including July 10, 2012, and July 11, 2012.

It is our understanding that your intentions are to terminate this license in the very near future.

However, there appears to have been a delay in gathering and transmitting the information to us that is needed in order to complete your license's termination before the renewal of your license would become "due" in our internal tracking system.

Therefore, we are issuing your renewed license for an abbreviated, one-year term only in order to permit you the time needed to prepare a high quality termination application. As such, it will not expire until July 31, 2013.

Please note that the major changes made to your license are printed in bold font.

As you know, we cannot authorize licensees to release the locations of use from licenses for unrestricted use (even by other members of their staff) until we have received and reviewed a copy of the results of decommissioning and close-out surveys for the facilities.

The final status survey must include a complete historical review of all actual licensed materials used, including sealed and unsealed sources, spills, and contamination. If sealed sources were transferred as part of the close-out of this license, please provide a copy of the final leak test for each sealed source; a copy of an acknowledgment of receipt from the licensed entity who took possession of each source; and if the recipient/transferee is an Agreement State license, please include a current copy of its license that clearly shows it is licensed to receive your sources.

R. DeWeese

These references may also assist you: 10 CFR 30.41; 10 CFR 30.51; NUREG 1556 Vol. 9, Rev. 2, section 11, "Termination of Activities," and Appendix H, "NRC Form 314."

Your complete historical review should specify when and where all licensed materials, including materials in 10 CFR 31.11, were used, when the last use was for each authorized modality and how, when and by whom were the materials disposed of (shipped off site, decayed -in-storage, sanitary sewer disposal, etc.) or transferred.

Please respond by stating exactly which licensed materials were used at this location historically and please submit final status survey information covering those radioactive materials.

The final status survey should consist of exposure rate measurements to show that all sources of radioactive material have been removed, and contamination checks of areas where radioactive materials were used or stored. Average radiation levels associated with surface contamination and removable contamination should not exceed those specified in the decontamination guide being sent to you under separate cover.

Please also refer to section 15.5.3 in NUREG 1757, Vol. 1, Rev. 1, available on our website, for additional assistance.

Please submit the following information with your close-out survey:

- a. Diagrams of each facility with exposure rate survey and wipe test results keyed to specific locations, as appropriate.**
- b. The name of the person performing the survey.**
- c. The date the survey was performed.**
- d. The instrument(s) used for exposure rate measurements and for analysis of the wipes.**
- e. Background readings and each instruments' efficiency or correction factor.**
- f. The date(s) that the survey instruments were last calibrated.**
- g. The action levels for both exposure rate measurements and wipe tests. Include the identity of areas exceeding these levels, corrective actions taken and results of corrective actions taken.**

Also, please always include the telephone number and fax number of at least one person who serves as a point of contact for all future licensing requests. Ensure that a management representative signs the amendment request, in accordance with 10 CFR 35.12(a).

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region III office at (630) 829-9887 so that we can provide appropriate corrections and answers.

R. DeWeese

NRC's Regulatory Issue Summary (RIS) 2005-31 provides criteria to identify security-related sensitive information and guidance for handling and marking of such documents. This ensures that potentially sensitive information is not made publicly available through ADAMS, the NRC's electronic document system.

Pursuant to NRC's RIS 2005-31 and in accordance with 10 CFR 2.390, the enclosed license document is exempt from public disclosure because its disclosure to unauthorized individuals could present a security vulnerability.

The RIS may be located on the NRC Web site at: <http://www.nrc.gov/reading-rm/doc-collections/gen-comm/reg-issues/2005/ri200531.pdf> and the link for frequently asked questions regarding protection of security related sensitive information may be located at: <http://www.nrc.gov/reading-rm/sensitive-info/faq.html>.

A copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions, and representations made in your license application and supplemental correspondence with NRC will result in enforcement action against you.

This could include issuance of a notice of violation, or imposition of a civil penalty, or an order suspending, modifying or revoking your license as specified in the General Statement of Policy and Procedure for NRC Enforcement Actions.

Since serious consequences to employees and the public can result from failure to comply with NRC requirements, prompt and vigorous enforcement action will be taken when dealing with licensees who do not achieve the necessary meticulous attention to detail and the high standard of compliance which NRC expects of its licensees.

The NRC's Safety Culture Policy Statement became effective in June 2011. While a policy statement and not a regulation, it sets forth the agency's *expectations* for individuals and organizations to establish and maintain a positive safety culture.

You can access the policy statement and supporting material that may benefit your organization on NRC's safety culture Web site at <http://www.nrc.gov/about-nrc/regulatory/enforcement/safety-culture.html>.

R. DeWeese

We strongly encourage you to review this material and adapt it to your particular needs in order to develop and maintain a positive safety culture as you engage in NRC-regulated activities.

Sincerely,

A handwritten signature in cursive script that reads "Colleen Carol Casey".

Colleen Carol Casey
Materials Licensing Branch

License No. 13-26797-01
Docket No. 030-34441

Enclosure:

Amendment No. 08