

TRANSMISSION VERIFICATION REPORT

3 tried Rhonda calling Lennon to leave a VM using but system didn't work - in directory, dial pad...

TIME : 06/26/2012 19:32
NAME : USNRC RIII
FAX : 6308299782
TEL :
SER.# : 000A7J925774

DATE, TIME
FAX NO./NAME
DURATION
PAGE(S)
RESULT
MODE

06/26 19:31
88123766665
00:01:00
05
OK
STANDARD
ECM

*7:49 pm
CT 6/26/12
CCEC*

NRC FORM 386 (RIII)



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4352

TELEFAX TRANSMITTAL

DATE:

6/26/12

NUMBER OF PAGES:
(including this page)

6

SEND TO:

Rhonda Lennon, Practice Manager

LOCATION:

Columbus Imaging Associates

FAX NUMBER:

812 - 376 - 6665

VERIFY BY CALLING SENDER

FROM:
(SENDER)

Colleen Carol Casey

TELEPHONE NUMBER:

630 - 829 - 7844

FAX NUMBER:

630 -

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE

Please call me to discuss this if you



UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION III
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4352

TELEFAX TRANSMITTAL

DATE: 6/26/12

NUMBER OF PAGES: 6
(including this page)

SEND TO: Rhonda Lennon, Practice Manager

LOCATION: Columbus Imaging Associates

FAX NUMBER: 812 - 376 - 6665 **VERIFY BY CALLING SENDER**

FROM: Colleen Carol Casey
(SENDER)

TELEPHONE NUMBER: 630 - 829 - 7841 FAX NUMBER: 630 - -

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE

Please call me to discuss this if you have any questions.

Thank you very much.

Colleen Carol Casey

NOTICE

This message is intended only for the use of the individual or entity to which it is addressed and may contain information that is privileged, confidential, or exempt from disclosure under applicable law. If the reader of this message is not the intended recipient or the employee responsible for delivering the message to the intended recipient, you are hereby notified that any dissemination, distribution or copying of this communication is strictly prohibited. If you have received this communication in error, please notify the sender immediately by telephone and return the original to the above address, by U.S. Mail. Thank you.

CONVERSATION RECORD

COLLEEN CAROL CASEY
MATERIALS LICENSING BRANCH
UNITED STATES NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD STE 210
LISLE, ILLINOIS 60532-4352
OFFICE: (630)-829-9841 FAX: (630) 515-1078
EMAIL: COLLEEN.CASEY@NRC.GOV

	TIME	DATE
ACTUALLY FAXED / EMAILED?	YES.	June 26, 2012

Fax No.: 812-376-6665

Email address: @ .

NAME OF PERSON(S) CONTACTED	ORGANIZATION	TELEPHONE NO.
Rhonda Lennon, Practice Manager for Columbus Imaging Associates		812-376-1000

SUBJECT

License No.: 13-26797-01

Control No.: 576969

SUMMARY

We have reviewed your letter dated January 25, 2012, including attached application dated February 9, 2012, requesting renewal of your byproduct materials license and find that we need additional information as follows:

1. Please always include the name of at least one knowledgeable point of contact, his/her telephone number, his/her fax number and his/her email address to facilitate communications between us.

In your letter above, we were not provided with a fax number for you or an email address. We used the fax number given on your letterhead as a default. Please advise us of a more appropriate fax number for you and/or Dr. Deweese and your email addresses, if it is acceptable for us to use them to more efficiently communicate with you.

2. Items 5, 6 and 7 of your application appear to request continued authorization for the radioactive materials and authorized users on this license. However, instead of referencing 10 CFR 35.100, 35.200 and 35.300, your application references 10 CFR 35.190, 35.290 and 35.390, which are the training and experience requirements for authorized users.

2

The references for 10 CFR 35.100, 35.200 and 35.300 should have been used. In response, please confirm that it was your intention to refer to “10 CFR 35.100, 35.200 and 35.300” and not “10 CFR 35.190, 35.290 and 35.390.”

Please consult “Consolidated Guidance About Materials Licenses: Program - Specific Guidance About Medical Use Licenses (NUREG-1556, Volume 9, Revision 2)”, Appendix C and Tables C.2 and C.3 for assistance.

Please also see a section on advice on preparing license renewals attached to this record. This section was added because your application contained significant amounts of information that we no longer require. If you had limited your submittal to only the information we requested in the NUREG, you would have had a more complete application and greatly reduced the regulatory burden on yourselves.

Here is a link to the NUREG 1556, Vol. 9, Rev.2:

<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>

3. Please note that we no longer exclude generators and xe-133 on medical licenses. Since we will be listing your application in Condition No. 16 of your license when your renewal is completed, these restrictions and exclusions will become part of your license in that manner. This also applies to restrictions and exclusions that may be requested for your authorized users in the future. This is a no response item.
4. Your application did not include a possession limit for iodine-131 in your authorization request for 10 CFR 35.300.

Please specify a possession limit for iodine-131 in 10 CFR 35.300, or Subitem Nos. 6 through 9 D of your license. Your possession limit must be less than ten curies, be realistic for your needs and include waste stream activities as well. Many licensees typically request one curie for this purpose, which is also the current possession limit for iodine-131 on your license.

5. Your application did not specify whether you are using, or plan to use, PET materials. Please advise us in response whether you are using or plan to use PET materials as significant additional information will be required to approve such a request.
6. Your application was silent with respect to continuing authorization for Drs. Robert R. Hasson, Richard Pitman, Kenneth Grant Scott and Michael L. Lutz as authorized users.

Please note that since you did not explicitly direct us to delete their authorizations, we will normally continue them as we consider such silence to be “an oversight.”

If you wish to have these four authorized users removed from your license, please so state explicitly in your response.

If you wish to continue these authorizations, please so state explicitly also.

3

Please also be reminded of the provisions in 10 CFR 35.14(b)(1) requiring that NRC be notified in writing within 30 days after an authorized user permanently discontinues duties under the license.

Please also be reminded of the provisions in 10 CFR 30.9(a), "Completeness and accuracy of information,"..."(a) Information provided to the Commission by an applicant for a license or by a licensee or information required by statute or by the Commission's regulations, orders, or license conditions to be maintained by the applicant or the licensee shall be complete and accurate in all material respects."


In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

ACTION REQUIRED

Submit the requested information within **15 calendar days (by July 12, 2012)** by referencing control number **576969** to facilitate proper handling in our office. Please contact me if you need to make alternative response arrangements.

Upon receipt of your response we will reactivate placement of your request in our database and resume our review. Address your written response, via an appropriately dated and signed (by management) cover letter, to my attention at the above address.

PLEASE DIRECT ANY QUESTIONS YOU MAY HAVE TO ME AT 630-829-9841 or 800-522-3025, ext. 9841.

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Colleen Carol Casey		June 26, 2012

Additional Guidance for Renewing an NRC License

Please note that using the NUREG 1556 series documents will help ensure that your licensing correspondence is prepared more completely and in a less burdensome manner.

In preparing future licensing correspondence please focus on providing the information requested in Appendix C to NUREG 1556, Volume 9, Rev. 2. Follow the "Suggested Format.." provided in this Appendix and use the suggested responses and model procedure/appendix references whenever possible, appending descriptive information as appropriate.

It is strongly advisable to read the corresponding text in the front of each NUREG to ensure a complete understanding of the commitments that you make.

Please do not submit resumes, curricula vitae, college transcripts, any personal, proprietary information, blueprint diagrams, or copies of blueprint diagram and any extraneous, prescriptive information and procedures, unless we specifically request it, which is unlikely.

If you must deviate from a model procedure or suggested response, it may be possible to simply indicate what the deviation is and still use the model procedure/ suggested response as a "basic" commitment.

Descriptive information may be "recycled" from previous documents only so long as it is current, complete information equivalent to the model procedure (as appropriate) and does not contain extraneous material.

It is in your best interests to only provide those commitments, statements, representations and procedures, in a clear and explicit manner, that we require to issue your license. These documents will form the basis for the license in the last condition of the license, called the "tie-down" condition.

You will realize the benefit of a reduced regulatory burden while enhancing safety and maintaining compliance and efficiency if you establish your license in this manner.

In fact, the easiest way to prepare a renewal, for example, is to take a copy of NUREG 1556, Vol. 9, Rev. 2, Appendix C, especially Tables C.2 and C.3 to your copy machine and copy it out directly. Read the text in the front of the NUREG that corresponds to each section and simply fill in the checkmarks and blanks on the copied checklist, thereby making your license commitments.

Please do not re-type the checklist as errors and omissions may be introduced.. As you need to append certain information or provide an alternative procedure, please be sure to incorporate the information in the NUREG, at a minimum, to ensure completeness.

Please see the Tables C.2 and C.3, NUREG 1556, Vol. 9, Rev. 2, Appendix C. A hard copy of this document should have been sent to you already. It is also available on our website at:

<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>

If you wish, you may contact me at the telephone number above to discuss this further . We would be happy to assist you as we are able.