

Conversation Record

FROM
TO: Daryl Moore

COMPANY: Xerium Technologies, Inc.

TO:
 Jennifer Bishop
 Material Licensing Branch
 U.S. Nuclear Regulatory Commission
 Region II
 2443 Wrenville Road
 Lisle, Illinois 60532-4351
 Jennifer.Dalzell-Bishop@nrc.gov

PAGES: 2 **TEL.:** 540-869-6111

FAX #: 540-526-9192

(630) 821-9607 **FAX: (630) 515-1078**

EMAIL: Daryl.Moore@xerium.com

CONVERSATION RECORD

TIME **DATE**

NAME OF PERSON(S) CONTACTED

ORGANIZATION

Daryl Moore

Radiation Safety Officer

SUBJECT

License No.: 45-31371-01

Control No.: 577891

SUMMARY

This refers to your letters to amend your license, dated July 12, 2012 and July 30, 2012, and the phone conversation between Daryl Moore and Jennifer Bishop on August 22, 2012.

In your letter, you requested that one location of use be removed from your license. In support of this request, please provide the following information.

RESPONSE: Please provide the most current leak test results from all of the gauges that were located at that site. If a gauge is not longer in your possession, please provide the documentation that this device has been properly transferred.

In addition, please confirm, in writing that there were no leaking sources stored at the location you requested to remove from the license.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically in the NRC Public Document Room or from the Publicly Available Record (PIRS) component of NRC's document system (ADAMS) accessible from the NRC Website at <http://www.nrc.gov/reading-rem/adams.html>.

Submit the requested information within the original timeframe of 7 calendar days (by August 29, 2012) by referencing control number 577891 to facilitate proper handling in our office. Please contact me if you need to make alternative response arrangements.

D. Moore | Xerium Technologies, Inc., Lic. No. 45-31371-01

p. 2 of 2

PLEASE DIRECT ANY QUESTIONS YOU MAY HAVE TO ME AT (630) 829-9607.

For future reference, please always include the name, phone number and fax number of at least one person whom we may contact for additional information when reviewing your licensing correspondence and requests. Providing an email address for the contact person is also very helpful.


NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE DATE

Jennifer Bishop



8/22/12



Xerium Technologies, Inc.
P.O. Box 575
Middletown, VA 22645
Phone: (540)-508-3044
Fax: (540)-526-9192

August 22, 2012

Ms. Jennifer Bishop
Materials Licensing Branch
U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Lisle, IL 60532-4351

Subject: Addition Information for Amendment Request License 45-31371

Dear Ms. Bishop:

We have attached for your review a copy of the most recent leak test for the portable gauge stored at our old site in Mattawan, MI.
There were no leaking sources stored at the Mattawan, MI site.

Please feel free to contact us for any additional information.

Sincerely,

Daryl Moore, RSO
Safety & Environmental Manager



STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.	97681		
XERIUM TECHNOLOGIES - MIDDLETOWN			
LOCATION	22645 - MIDDLETOWN		
CONTACT PERSON:	DARYL MOORE	PHONE 1-540-869-8111 X 106	
XERIUM TECHNOLOGIES MIDDLETOWN 8207 VALLEY PARK MIDDLETON, WI 53561 UNITED STATES OF AMERICA			
LEAK TEST ANALYSIS			
EQUIPMENT NO.	NOT AVAILABLE	DEVICE/TOOL NUMBER	NOT AVAILABLE
SOURCE SERIAL NO.	1195	MANUFACTURER S/N	
ISOTOPE	AM-241	SOURCE ORIGINAL ACTIVITY	25 mCi 0.925 GBq
DATE SOURCE WIPE D	11 Jun 2012	REASON FOR TEST	PERIODIC
ANALYSIS DATE	18 Jun 2012	ANALYTICAL METHOD	Liquid Scintillation Counting
RAW SAMPLE COUNT	27 cpm	INSTRUMENT	Beckman LS 6500 706-9054
BACKGROUND	25 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	2.06737 dpm	MINIMUM DETECTABLE ACTIVITY	1.24E-05 μ CI 0.46 Bq
SAMPLE ANALYSIS ACTIVITY	0 μ CI 0.03509 Bq	PASS/FAIL CRITERION	0.005 μ CI 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 μ CI 37 Bq
ANALYST	Keith Gilbert <i>Keith Gilbert</i>		
NOTES:			
ADDITIONAL SOURCE DESCRIPTION:	SHUTTER CHECKED		

PRINTED IN THE UNITED STATES OF AMERICA
510 COUNTY HIGHWAY V TWO RIVERS, WISCONSIN 54241 VOICE: +1. 920.686.3889 FAX: 1.920.686.3889