

**U.S. NUCLEAR REGULATORY COMMISSION**

Date:

**TELEPHONE CONVERSATION RECORD**

8-28-12

Mail Control  
or Report No(s).

578141

License No(s).

08-03075-07

Docket No(s).

03001321

Name of Licensee:

Howard University Hospital

Name of Participant(s):

Dr. Satya Bose

Telephone No.

202-806-7216

Subject:

RAI for therapy treatment room

(NOTE: This will be used as the  
Documents Title in ADAMS)

Discussed with Dr. Bose that the dose rates in the adjoining patient room exceeded 2 mR/hour. Dr. Bose indicated that he would review the shielding evaluation and the possibility of restricting access to the adjacent patient room during therapy treatments and submit additional information.

**Action Required: Await Reply**

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Sensitive – Privacy Act/ PII

Immediate Release

Non-Sensitive Copyright

Sensitive – Internal

Normal Release Date: 09/06/2012

Sensitive- Proprietary

Sensitive – Security-Related

Delayed Release Date:

SUNSI Review Completed By:

PL

/ RA /