

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 8/31/2012

**TELEPHONE CONVERSATION RECORD**Mail Control  
or Report No(s).

577216

License No(s).

01-07317-01

Docket No(s).

03003576

Name of Licensee:

U.S. EPA

Name of Participant(s):

John Kirby, RSO  
Dennis Lawyer, R1, DNMS

Telephone No.

334-270-3439

Subject:

RAI, Sealed Sources

(NOTE: This will be used as the  
Documents Title in ADAMS)

I called Mr. Kirby left a message and he returned my call. I discussed the items submitted in letter dated August 14, 2012. I told him that I would take the Certification of Financial Assurance as a request to change the amount of Byproduct material (as defined in 10 CFR 40.4 on his license. I stated that the only remaining item on my "Request for Further Information" was the sealed source numbers for Co-60 and Ni-63. Mr. Kirby said that he did not have sealed source numbers for Co-60 and that it could be licensed under his authorization in 6.A, 7.A, and 8.A. He then discussed that the model number Agilent G2397A and G1223A were used. I told them that these were generally licensed devices and could not be licensed under the specific license. I then discussed general license requirements as written in 10 CFR 31.5. He said to remove the Ni-63 authorization.

**Action Required: Proceed with License Renewal**

Document Availability:

 Publicly Available Non-Publicly Available Non-Sensitive  
 Sensitive – Privacy Act/ PII  
 Immediate Release Non-Sensitive Copyright  
 Sensitive – Internal  
Normal Release Date: 09/11/2012 Sensitive- Proprietary  
 Sensitive – Security-Related  
Delayed Release Date:SUNSI Review Completed By: Dennis Lawyer / RA /