

**From:** [Weidner, Tara](#)  
**To:** "[melleck88@gmail.com](mailto:melleck88@gmail.com)"  
**Subject:** NRC-Request for additional information  
**Date:** Tuesday, September 04, 2012 10:36:00 AM

---

License No. 08-31176-01  
Docket No. 030-37314  
Control No. 578054

**SUBJECT: REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION FOR AMENDMENT TO LICENSE, CONTROL NO. 577948**

**\*\*\*\*\* PLEASE RESPOND TO CONFIRM RECEIPT OF THIS REQUEST \*\*\*\*\***

Dear Dr. Malik:

This is in reference to your letter dated August 8, 2012 requesting to amend Nuclear Regulatory Commission License No. 08-31176-01. In order to continue our review, we need the following additional information:

1. On your facility diagram you did not indicate what is adjacent to and above/below the new facility. Please specify.
2. It appears that the hot lab is located within the camera room. Please explain how you will maintain security of the byproduct materials stored in the camera room.
3. Confirm that your use of byproduct material at the new location will be the same as at your current location.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 578054. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5272.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

*Tara L. Weidner*  
*Senior Health Physicist*  
*US Nuclear Regulatory Commission*