

Roldan, Lizette

From: Roldan, Lizette
Sent: Friday, August 31, 2012 2:43 PM
To: 'gowens@amengtest.com'
Subject: REQUEST FOR ADDITIONAL INFORMATION REGARDING RENEWAL REQUEST CONTROL 577260
Attachments: sr1556v1r1.pdf; apph-errata.pdf

License No.: 22-20271-01
Docket No.: 030-34780
Control No.: 577260

Dear Mr. Owens:

This is in reference to your application dated March 7, 2012 requesting to renew Nuclear Regulatory Commission License No. 22-20271-01. In order to continue our review, we need the following additional information:

1. Please provide a statement: "We will either possess and use, or have access to and use, a radiation survey meter that meets the Criteria in section entitled "Radiation Safety Program – Instruments" in NUREG-1556, Vol. 1, Rev. 1, dated November 2001."
2. Please provide a statement: "Physical inventories will be conducted at intervals not to exceed 6 months, to account for all sealed sources and devices received and possessed under the license."
3. Please provide a statement: "We will implement and maintain the operating, emergency and security procedures in Errata: Appendix H of NUREG-1556, Vol. 1, Rev. 1, dated November 2011, and provide copies of these procedures to all gauge users and at each job site."
4. Please provide a statement: "We will send the gauge to the manufacturer or other person authorized by NRC or an Agreement State to perform non-routine maintenance or repair operations that require the removal of the source or source rod from the gauge."

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Regulations, Guidance, and Communications**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 7:00 a.m. to 6:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region IV Office and refer to Mail Control No. 577260. If you have any technical questions regarding this deficiency letter, please call me at (817) 200-1596.

If we do not receive a reply from you by **September 14, 2012**, we will assume you do not wish to continue with the renewal process.

Sincerely,

Lizette Roldán-Otero, Ph.D.

Health Physicist
U.S. Nuclear Regulatory Commission
1600 E. Lamar Blvd.
Arlington, TX 76011-4511

Office: 817-200-1596
Fax: 817-200-1188