

From: [Elliott, Robin](mailto:Robin.Elliott@nrc.gov)
To: "carwer41@hotmail.com"
Subject: Renewal of License No. 52-16061-01
Date: Thursday, September 06, 2012 6:27:00 PM

Licensee: Saint Luke's Memorial Hospital, Inc.
Licensee Number: 52-16061-01
Docket Number: 03010302
Mail Control Number: 577239

To: Carmelo Pérez, Radiation Safety Officer

This is in reference to the renewal application for this license. Thank you for your response to the previous request for information. In order to complete my review, I need to clarify the following information as we discussed on the phone today.

Please send a return email to confirm that you received this message.

1. You indicated in your last communication that the following radiation monitoring equipment will be used to perform radiation level, detection, measurement and removable contamination surveys for your license: LUDLUM M-14C and an uptake System well counter CAPINTEC Model CAPTUS-600 Spectrometer with a detector well type NaI (TI) 2" x2" crystal. Please clarify if this is not what you intended.
2. Please provide leak test certificates for the sources transferred from the Calle Guadalupe Final location to the St. Luke's Memorial Hospital Machuelo location. Please note that per 10 CFR 35.67 if the activity of the sources was 100 microcuries or less for beta or gamma-emitting material or 10 microcuries or less of alpha-emitting material, no leak test is required.

Please provide a written response to these items as soon as possible but no later than September 13, 2012 under signature of hospital senior management. This may be provided to my attention by letter or fax (610-337-5269), referencing mail control number 577239.

If you respond by letter, the new mailing address for NRC Region 1 is 2100 Renaissance Blvd., Suite 100, King of Prussia, PA 19406-2713.

Please contact me by telephone or email with any questions.

Regards,

Robin L. Elliott

Health Physicist
U. S. Nuclear Regulatory Commission
Region 1
2100 Renaissance Blvd, Suite 100
King of Prussia, PA 19406-2713
(610) 337-5076 voice
(610) 337-5269 fax

robin.elliott@nrc.gov