

From: Simmons, Michelle
Sent: Tuesday, September 11, 2012 2:41 PM
To: Kketchum@peacehealth.org
Subject: Peace Health Ketchikan Medical Center Deficiency Letter

Docket: 030-19521
License: 50-19913-01
Control: 577875

This is in reference to your application dated July 9, 2012, requesting to renew your material license to Nuclear Regulatory Commission License No. 50-19913-01. In order to continue our review, we need the following additional information:

Please confirm the following statements:

- "Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations."
- "We have developed and will implement and maintain written procedures for area surveys in accordance with 10 CFR 20. 1101 that meet the requirements of 10 CFR 20.1501 and 10 CFR 35.70."

Please submit the additional information by September 21, 2012. If you are unable to respond by this due date, please don't hesitate to contact me so we can discuss an extension to the date. Our fax number is (817) 200-1263. You may also email your response as a pdf attachment. If you have any questions regarding this email, please call me at 817-200-1590. When responding to this email, please include the license, docket and control numbers located at the top of this page.

Thanking you in advance for your cooperation, assistance, and prompt response in this matter.

Michelle Simmons | Health Physicist | U.S. Nuclear Regulatory Commission

☎ (817) 200-1590 | ✉ michelle.simmons@nrc.gov

PUBLIC

- Immediate Release
 Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other: _____

Reviewer: CA Date: 9/11/12