

**From:** [Nguyen, Janice](mailto:Nguyen_Janice)  
**To:** ["kenneth.wohlt@cvmc.org"](mailto:kenneth.wohlt@cvmc.org)  
**Subject:** NRC Request for Information - Mail Control Number 577696  
**Date:** Wednesday, September 12, 2012 11:16:00 AM

---

Licensee: Central Vermont Medical Center  
License Number: 44-13353-01  
Docket Number: 030-03292  
Mail Control Number: 577696

Dear Mr. Wohlt,

Could you please reply back to this email to confirm receipt?

This is in reference to your renewal application dated May 30, 2012, requesting to renew Nuclear Regulatory Commission License No. 44-13353-01.

In order to continue our review, we need the following additional information:

1. Please confirm that the mailing address for the license is Barre, Vermont.
2. Please confirm that you are not using PET materials.
3. Please resubmit your facility diagrams, showing the licensed areas of use in a more close-up view. Please clearly delineate rooms where radioactive material will be used, including all adjacent areas and areas above and below, and indicate their designated use. Please also include diagrams of Suite 2-1, that was previously licensed under Central Vermont Cardiology Associates.
4. Please provide facility diagrams of the rooms where prostate seed patients will be housed if they cannot be released under 10 CFR 35.75, including areas above, beside, and below. Otherwise, you may confirm that you will only implant patients with I-125 and Pd-103 only, who are releasable under 35.75.
5. Please confirm that "We have developed and will implement and maintain written procedures for area surveys in accordance with 10 CFR 20.1101 that meet the requirements of 10 CFR 20.1501 and 10 CFR 35.70."
6. Please confirm that "We have developed and will implement and maintain written waste disposal procedures for licensed material in accordance with 10 CFR 20.1101, that also meet the requirements of the applicable section of 10 CFR Part 20, Subpart K, and of 10 CFR 35.92."
7. Please confirm the model number of your Ludlum 14C survey meter probes.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

Please have your response signed by a member of senior management. You may respond to my attention in writing by letter, email (if letter is scanned into a pdf format), or fax (610-337-5269), referencing mail control number 577696.

If we do not receive a reply from you within 30 days, we will assume that you do not wish to pursue your application. Please feel free to contact me with any questions you may have.

Please note that the office of the Region I USNRC Division of Nuclear Materials Safety has moved effective May 9, 2012. Our new address is:

U. S. Nuclear Regulatory Commission  
Region I  
2100 Renaissance Blvd, Suite 100  
King of Prussia, PA 19406-2713

Thank you in advance for your help.

*Jan Nguyen*

Janice Nguyen  
Health Physicist  
US Nuclear Regulatory Commission Region I  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406  
(610) 337-5006  
FAX (610) 337-5269