

U.S. NUCLEAR REGULATORY COMMISSION

Date: 09/18/12

TELEPHONE CONVERSATION RECORD

Time: 09:20am

Mail Control
or Report No(s).

577810

License No(s). 06-31418-01

Docket No(s). 03038340

Name of Licensee: Connecticut Oncology & Hematology

Name of Participant(s): Lisa Shomsky, Business Administrator

Telephone No. (860) 482-5384 ext. 230

Subject: Additional information for termination request

(NOTE: This will be used as the
Documents Title in ADAMS)

NRC staff contacted the licensee to clarify the following information provided in a deficiency response dated September 12, 2012:

1. Verification that the licensee is no longer in operation. Lisa verified that they have not provided any PET services since June 15, 2012.
2. The leak test record provided did not include readings for three Germanium-68 rod sources. I asked Lisa to provide the leak tests for these.
3. The Cs-137 source (Model GF-0008, SN 1181-63-82) listed on the leak test and the inventory records showed different activities (0.5 mCi and 0.5 μ Ci). I asked Lisa to verify the actual activity of the source.

Lisa agreed to provide this information by COB today.

Action Required: Issue termination

Document Availability:

 Publicly Available Non-Publicly Available
 Non-Sensitive
 Sensitive – Privacy Act/ PII
 Immediate Release

 Non-Sensitive Copyright
 Sensitive – Internal
 Normal Release
 Date:09/26/2012

 Sensitive- Proprietary
 Sensitive – Security-Related
 Delayed Release Date:

SUNSI Review Completed By: M. Abogunde / RA /

Document Accession #: _____