

**From:** [Nguyen, Janice](#)  
**To:** [Razmianfar, Nasser](#)  
**Cc:** [jhack@hsc.wvu.edu](mailto:jhack@hsc.wvu.edu)  
**Subject:** HDR Relocation  
**Date:** Tuesday, October 09, 2012 11:49:00 AM

---

West Virginia University Hospital, Inc.  
Mail Control Number 577889  
License Number 47-23066-02  
Docket Number 030-20233

Dear Mr. Razmianfar,

Could you please reply back to this email to confirm receipt?

This is in reference to your letter dated July 13, 2012, requesting to amend Nuclear Regulatory Commission License No. 06-01060-01.

In order to continue our review, we need the following additional information with regard to the revised HDR spot check procedures:

Please provide a letter from the electrician who installed the door interlock switch and the HDR/gamma knife switch. He should provide the circuit schematics and confirm that the door switch selector is in series with the door contact so that when either the door is opened during treatment or when the door switch selector is changed from HDR to gamma knife, it results in a door open status and source retraction. Please also confirm that you have tested the door interlock by opening the door during treatment and it resulted in source retraction.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

You may respond to my attention in writing by letter, email (if letter is scanned into a pdf format), or fax (610-337-5269), referencing mail control number 577889.

If we do not receive a reply from you by October 12, we will assume that you do not wish to pursue your amendment. Please feel free to contact me with any questions you may have.

Please note that the office of the Region I USNRC Division of Nuclear Materials Safety has moved effective May 9, 2012. Our new address is:

U. S. Nuclear Regulatory Commission  
Region I  
2100 Renaissance Blvd, Suite 100  
King of Prussia, PA 19406-2713

Thank you in advance for your help.

*Jan Nguyen*

Janice Nguyen  
Health Physicist  
US Nuclear Regulatory Commission Region I  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406  
(610) 337-5006  
FAX (610) 337-5269