

David Rhoe
Pasoe de la Fuente
D4 Calle Tivoli
San Juan, PR 00926
787-245-7248

October 10, 2012


Lester Tripp
USNRC Region I DNMS
2100 Renaissance Blvd
King of Prussia, PA 19406

Control No. 579099

Dear Mr. Tripp:

This letter is to confirm that all licensed radioactive materials were removed from the facility. During the close out survey dated August 28, 2012, a visual inspection was performed to verify and account for all radiation sources. I personally transferred all licensed radiative materials were transferred to Fajardo facility. If you have any questions regarding this matter, please feel free to contact me at the number above.

Sincerely,



David Rhoe
Consultant

Leak Test Record
NRC License 52-25430-01

Leak Tested For: Caguas Nuclear Imaging Center
Leak Tested By: David Rhoe
Standard Source: Cs-137 NES-139S
Standard Activity: 0.105 uCi Nominally
Date of Standard: 9-Sep-88
Date of Leak Test: 7/25/2012
Decay Activity: 0.06046 From decay chart
Standard (dpm): 134221

Instrument: Gamma
Beckman Gamma
Instrument Model Number: 5500
Instrument Serial Number: 8044788
Standard (cpm) 34322
Efficiency: 0.25571 25.6 %
Counting time (minutes) 1
Background (cpm) 73
Minimum Detectable Activity: 2.128E-05

Wipe (Smear) Test: All external or accessible surfaces of the source or housing are wiped with a piece of filter paper or other absorbent material which has been moistened with an appropriate solvent and the activity removed is measured.

Source ID and Serial Number	Gamma	
	Wipe Test	Sample Activity
Cs-137 Sn103594 191.2 uCi 2/1/2007	72	0.00013
Ba-133 Sn103179 272.5 uCi 2/1/2007	75	0.00013
Co-57 Sn103142 5.557 mCi 2/1/2007	72	0.00013

This test reveals that 0.005 microcuries or less was present as removable contamination. Should the removable contamination exceed 0.005 microcuries, the source must be removed from use and necessary measures taken according to NRC regulations.


David Rhoe Health/Medical Physicist