

APPENDIX E

Instrument and Check Source Calibration Certificates



CERTIFICATE OF CALIBRATION

CUSTOMER US ARMY CORPS OF ENGINEERS ORDER NO. 20200290/378285
Mfg. Ludlum Measurements, Inc. Model 2360 Serial No. 138251
Mfg. Ludlum Measurements, Inc. Model 43-37-1 Serial No. PR136361
Cal. Date 23-May-12 Cal Due Date 23-May-13 Cal. Interval 1 Year Meterface 202-855

Check mark applies to applicable instr. and/or detector IAW mfg. spec. T. 73 °F RH 36 % Alt 699.8 mm Hg
 New Instrument Instrument Received Within Toler. +10% 10-20% Out of Tol. Requiring Repair Other-See comments
 Mechanical ck. Meter Zeroed Background Subtract Input Sens. Linearity
 F/S Resp. ck. Reset ck. Window Operation Geotropism
 Audio ck. Alarm Setting ck. Batt. ck. (Min. Volt) 2.2 VDC RS-232 Port OK
 Calibrated in accordance with LMI SOP 14.8 rev 12/05/89. Calibrated in accordance with LMI SOP 14.9 rev 02/07/97.

Instrument Volt Set 1600 V
 HV Readout (2 points) Ref./Inst. 500 / 505 V Ref./Inst. 2000 / 2002 V

Firmware Version: 39010-25 (EEPROM Settings)
Alpha Threshold: 90 mv User Time: 1.0
Beta Threshold: 4 mv Alpha Alarm: 999999
Beta Window: 30 mv Beta Alarm: 999999
Overload Checked Bot Not Set. A/B Alarm: 999999
Instrument calibrated with a 39' cable. Model 2360 Date: 5/23/2012
High voltage set with detector NOT connected. Calibration Date Due: 5/23/2013

COMMENTS:

Pu239 SN:7053 Size:12600cpm, Background:6cpm, Counts:5281cpm, 2pi Eff:41.86%
Tc99 SN:5280 Size:58300cpm, Background:1002cpm, Counts:26891cpm, 2pi Eff:44.40%
SrY90 SN:5281 Size:69345cpm, Background:1002cpm, Counts:25826cpm, 2pi Eff:35.79%

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING*
x1000	400k cpm	400	400
x1000	100k cpm	100	100
x100	40k cpm	400	400
x100	10k cpm	100	100
x10	4k cpm	400	400
x10	1k cpm	100	100
x1	400 cpm	400	400
x1	100 cpm	100	100

*Uncertainty within ± 10% C.F. within ± 20%

ALL Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*
Digital Readout					
400kcpm	39986(0)	39986(0)			
40kcpm	3998	3998			
4kcpm	400	400			
400cpm	40	40			
40cpm	4	4			

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements of ANSI/NCSL Z540-1-1994 and ANSI N323-1978 State of Texas Calibration License No. LO-1963

Reference Instruments and/or Sources: 73410 1131 781 059 280 60646 70897 Ra-226 S/N Y982
Cs-137 Gamma S/N 1162 G112 M565 5105 T1008 T879 E552 E551 720 734 1616 Neutron Am-241 Be S/N T-304
 Alpha S/N Beta S/N Other
 m 500 S/N 190566 Oscilloscope S/N Multimeter S/N 86250390

Calibrated By: Jana Flu Date 23-MAY-12

Reviewed By: Mark H Date 24 May 12



**CALIBRATION
CERTIFICATE**

EnergySolutions Services, Inc.
1570 Bear Creek Road
Oak Ridge, TN 37830
Phone: (877) 462-4873
Fax: (865) 220-4346

This Certificate will be accompanied by Calibration Charts or Readings where applicable

CUSTOMER INFORMATION				INSTRUMENT INFORMATION				
Customer Name: EnergySolutions Services, Inc.				Manufacturer: Ludlum				
Address: 1570 Bear Creek Road Oak Ridge, TN 37830				Model: 2360	Serial Number: 275713			
Contact Name: Tony Riggs				Probe: N/A	Serial Number: N/A			
Customer Purchase Order Number: N/A	Work Order 2011-11061 2011-11061 ²⁰¹¹⁻¹¹⁰⁶¹			Calibration Method: Electronic				
INSTRUMENT CALIBRATION INFORMATION								
Instrument Range	Calibration Standard Value	Ratemeter Response ($\pm 10\%$ of Standard Values)		Calibration Standard Value CPM	Time Base (minutes)	Tolerances (cpm) $\pm 2\%$	Scaler Response	
		As Found	As Left				As Found	As Left
X 1	100	100	100	1,000 CPM	0.1	90 - 110	100	100
X 1	250	250	250	1,000 CPM	0.5	450 - 550	500	500
X 1	400	400	400	1,000 CPM	1	900 - 1,100	1,002	1,002
X 10	1,000	1,000	1,000	1,000 CPM	2	1.8K-2.2K	2,003	2,003
X 10	2,500	2,500	2,500	1,000 CPM	5	4.5K-5.5K	5,008	5,008
X 10	4,000	4,000	4,000	1,000 CPM	10	9K-11K	10,015	10,015
X 100	10,000	10,000	10,000					
X 100	25,000	25,000	25,000					
X 100	40,000	40,000	40,000					
X 1000	100,000	100,000	100,000	Calibrated in accordance with OEM Technical Manual				
X 1000	250,000	250,000	250,000					
X 1000	400,000	400,000	400,000					
STATEMENT OF CERTIFICATION								
We Certify that the instrument listed above was calibrated and inspected prior to shipment and that it met all the Manufacturers published operating specifications. We further certify that our Calibration Measurements are traceable to the National Institute of Standards and Technology. (We are not responsible for damage incurred during shipment or use of this instrument).								
Instrument								
Calibrated By: <i>M. Paul</i>				Reviewed By: <i>J. Outen</i>		Date: 8/30/11		
Calibration Date: 08/30/2011				*Calibration Due (6mo): 02/29/2012				
				*Calibration Due (12mo): 08/30/2012				

* Calibration due date is dependant on users regulatory requirements.

Model: 2360

Serial Number: 275713

M&TE				Environmental Conditions		
Volt Meter	ID# 94710023	Cal Due: 10/28/2011		Barometer	ID# 3590	Cal Due: 09/21/11
Pulser	ID# 112860	Cal Due: 04/26/2012		Thermometer	ID# 3590	Cal Due: 09/21/11
Humidity	ID# 958670	Cal Due: 06/07/2012		Temp: 22.4 °C	Pressure: 741 mmHg	Humidity: 58%
Special Test						
BAT Check		Sat (✓) Unsat ()		Geotropism		Sat (✓) Unsat ()
LCD Display Check		Sat (✓) Unsat ()		Audio Check		Sat (✓) Unsat ()
Mechanical Zero		Sat (✓) Unsat ()		Low BAT Set		Sat (✓) Unsat ()
Reset		Sat (✓) Unsat ()				
HV Analog Display		Sat (✓) Unsat ()		As Found		As Left
High Voltage Calibration (± 10%)				Alpha Sensitivity = 125 mv	Alpha Sensitivity = 120 mv	
Voltage	Tolerance	As Found	As Left	Beta Sensitivity = 3.8 mv	Beta Sensitivity = 3.5 mv	
500	450-550	514	514	Beta Window = 32 mv	Beta Window = 30 mv	
1000	900-1100	983	983	Beta Setpoints--Pulser counts detected at 3.5mv ± 1mv and shut off at 30mv for beta. For Alpha channel counts detected at 120mv and above.		
1500	1350-1650	1,490	1,490			
H.V. Set With Detector Not Connected				Overload to be set with detector to be used		
COMMENTS						
<p>Calibrated in accordance with OEM Technical Manual</p> <p>See detector certificate for High Voltage setting</p> <p>**Calibrated with 5ft cable**</p>						
Instrument				Reviewed By: <i>J. Pulcinella</i> Date: 8/30/11		
Calibrated By: <i>M. Paul</i>				*Calibration Due (6mo): 02/29/2012		
Calibration Date: 08/30/2011				*Calibration Due (12mo): 08/30/2012		

* Calibration due date is dependant on users regulatory requirements.



**CALIBRATION
CERTIFICATE**

EnergySolutions Instrument Services
 1570 Bear Creek Road
 Oak Ridge, TN 37830
 Phone: (877) 462-4873
 Email: ISFstaff@energysolutions.com

This Certificate will be accompanied by Calibration Charts or Readings where applicable

CUSTOMER INFORMATION			DETECTOR INFORMATION		
Customer Name: EnergySolutions Instrument Services			Manufacturer: Ludlum		
Address: 1570 Bear Creek Road Oak Ridge, TN 37830			Detector Model: 43-37A		
Contact Name: Tony Riggs			Serial Number: 093966		
Customer Purchase Order Number: N/A		Work Order Number: 2011-11248	Evaluation Method: Source		
DETECTOR EFFICIENCY/RESPONSE/PRECISION INFORMATION					
Source Nuclide: Th ²³⁰	Serial Number: 119738	Activity: 18,600 dpm	2 Pi Emissions: 8,640	Certification Date: 10/20/97	
Parameter	As Found	As Left	Precision Test		CPM
Count 1	3,852	3,852	Count 1 (Heel)		3,514
Count 2	3,816	3,816	Count 2 (Center)		3,681
Count 3	3,652	3,652	Count 3 (Toe)		3,746
Count 4	3,773	3,773	Average		3,647
Count 5	3,594	3,594	Tolerance		±10%
Count 6	3,523	3,523	Pass/Fail		Pass
Average	3,702	3,702			
Background (CPM)	14.6	14.6			
Net Counts	3,687	3,687			
2pi Efficiency	42.7%	42.7%			
4pi Efficiency	19.8%	19.8%			
Low Sample Activity: Source #: N/A		High Sample Activity: Source #: N/A	Dead Time (DT): N/A	Calibration Constant (CC): N/A	
SCALER INFORMATION			OPERATING INFORMATION		
Model	Serial Number	Due Date	Background (cpm)	Operating Voltage	Threshold
2360	275713	08/30/2012	14.6	1700V	Alpha (120mV) Beta (3.5-30mV)
Detector Setup Report YES NO <input checked="" type="checkbox"/>		Barcode Report YES NO <input checked="" type="checkbox"/>		Voltage Plateau YES <input checked="" type="checkbox"/> NO	
COMMENTS					
5 minute background performed Efficiency performed on contact with 5Ft. cable 1 layer mylar (0.4mg/cm2)					
STATEMENT OF CERTIFICATION					
We Certify that the detector listed above was evaluated for proper operation prior to shipment and that it met all the Manufacturers published operating specifications. We further certify that our Calibration Measurements are traceable to the National Institute of Standards and Technology. (We are not responsible for damage incurred during shipment or use of this detector).					
Detector					
Certified By: M. Paul	Reviewed By: J. Dickinson	Date: 5/30/12			
Certification Date: 05/30/2012		* Certification Due (6mo): 11/30/2012			
		* Certification Due (12mo): 05/30/2013			

* Calibration due date is dependant on users regulatory requirements.

EnergySolutions Instrument Services
 1570 Bear Creek Road
 Oak Ridge, TN 37830
 Phone: (877) 462-4873
 Email: ISFstaff@energysolutions.com

This Certificate will be accompanied by Calibration Charts or Readings where applicable

CUSTOMER INFORMATION			DETECTOR INFORMATION		
Customer Name: EnergySolutions Instrument Services			Manufacturer: Ludlum		
Address: 1570 Bear Creek Road Oak Ridge, TN 37830			Detector Model: 43-37B		
Contact Name: Tony Riggs			Serial Number: 093966		
Customer Purchase Order Number: N/A		Work Order Number: 2012-12264	Evaluation Method: Source		
DETECTOR EFFICIENCY/RESPONSE/PRECISION INFORMATION					
Source Nuclide: Tc ⁹⁹	Serial Number: 099608	Activity: 21,312 dpm	2 Pi Emissions: 10,500	Certification Date: 08/08/96	
Parameter	As Found	As Left	Precision Test		CPM
Count 1	4,539	4,539	Count 1 (Heel)		4,553
Count 2	4,468	4,468	Count 2 (Center)		4,520
Count 3	4,429	4,429	Count 3 (Toe)		4,726
Count 4	4,493	4,493	Average		4,600
Count 5	4,765	4,765	Tolerance		±10%
Count 6	4,823	4,823	Pass/Fail		Pass
Average	4,586	4,586			
Background (CPM)	714.6	714.6			
Net Counts	3,872	3,872	Calibrated in accordance with CE-IN-WI-239		
2pi Efficiency	36.9%	36.9%			
4pi Efficiency	18.2%	18.2%			
Low Sample Activity: Source #: N/A		High Sample Activity: Source #: N/A		Dead Time (DT): N/A	Calibration Constant (CC): N/A
SCALER INFORMATION			DETECTOR INFORMATION		
Model	Serial Number	Due Date	Background (cpm)	Operating Voltage	Threshold
2360	275713	08/30/2012	714.6	1700 V	Alpha (120mV) Beta (3.5-30mV)
Detector Setup Report YES NO <input checked="" type="checkbox"/>		Barcode Report YES NO <input checked="" type="checkbox"/>		Voltage Plateau YES <input checked="" type="checkbox"/> NO	
COMMENTS					
5 minute background performed Efficiency performed on contact with 5Ft. cable 1 layer mylar (0.4mg/cm2)					
STATEMENT OF CERTIFICATION					
We Certify that the detector listed above was evaluated for proper operation prior to shipment and that it met all the Manufacturers published operating specifications. We further certify that our Calibration Measurements are traceable to the National Institute of Standards and Technology. (We are not responsible for damage incurred during shipment or use of this detector).					
Detector					
Certified By: <i>M. Paul</i>		Reviewed By: <i>Jeff Dikemiso</i>		Date: 5/30/12	
Certification Date: 05/30/2012			*Certification Due (6mo): 11/30/2012		
			* Certification Due (12mo): 05/30/2013		

* Calibration due date is dependant on users regulatory requirements.



**CALIBRATION
CERTIFICATE**

EnergySolutions Instrument Services
 1570 Bear Creek Road
 Oak Ridge, TN 37830
 Phone: (877) 462-4873
 Email: ISFstaff@energysolutions.com

This Certificate will be accompanied by Calibration Charts or Readings where applicable

CUSTOMER INFORMATION			DETECTOR INFORMATION		
Customer Name: EnergySolutions Instrument Services			Manufacturer: Ludlum		
Address: 1570 Bear Creek Road Oak Ridge, TN 37830			Detector Model: 43-37B		
Contact Name: Tony Riggs			Serial Number: 093966		
Customer Purchase Order Number: N/A		Work Order Number: 2012-12264	Evaluation Method: Source		
DETECTOR EFFICIENCY/RESPONSE PRECISION INFORMATION					
Source Nuclide: C ¹⁴	Serial Number: 010002		Activity: 260,460 dpm	2 Pi Emissions: N/A	Certification Date:
Parameter	As Found	As Left	Precision Test		CPM
Count 1	25,917	25,917	Count 1 (Heel)		25,671
Count 2	25,604	25,604	Count 2 (Center)		26,628
Count 3	26,697	26,697	Count 3 (Toe)		28,323
Count 4	25,882	25,882	Average		26,874
Count 5	28,245	28,245	Tolerance		±10%
Count 6	27,953	27,953	Pass/Fail		Pass
Average	26,716	26,716			
Background (CPM)	714.6	714.6			
Net Counts	26,002	26,002			
2pi Efficiency	N/A	N/A			
4pi Efficiency	10.0%	10.0%			
Low Sample Activity: Source #: N/A		High Sample Activity: Source #: N/A		Dead Time (DT): N/A	Calibration Constant (CC): N/A
SCALER INFORMATION			DETECTOR INFORMATION		
<u>Model</u>	<u>Serial Number</u>	<u>Due Date</u>	<u>Background (cpm)</u>	<u>Operating Voltage</u>	<u>Threshold</u>
2360	275713	08/30/2012	714.6	1700 V	Alpha (120mV) Beta (3.5-30mV)
Detector Setup Report YES NO <input checked="" type="checkbox"/>		Barcode Report YES NO <input checked="" type="checkbox"/>		Voltage Plateau YES <input checked="" type="checkbox"/> NO	
COMMENTS					
5 minute background performed Efficiency performed on contact with 5Ft. cable 1 layer mylar (0.4mg/cm2)					
STATEMENT OF CERTIFICATION					
We Certify that the detector listed above was evaluated for proper operation prior to shipment and that it met all the Manufacturers published operating specifications. We further certify that our Calibration Measurements are traceable to the National Institute of Standards and Technology. (We are not responsible for damage incurred during shipment or use of this detector).					
Detector					
Certified By: <i>M. Paul</i>		Reviewed By: <i>Jeff Dubins</i>		Date: 5/31/12	
Certification Date: 05/30/2012			*Certification Due (6mo): 11/30/2012		
			* Certification Due (12mo): 05/30/2013		



* Calibration due date is dependant on users regulatory requirements.



**CALIBRATION
CERTIFICATE**

EnergySolutions Services, Inc.
1570 Bear Creek Road
Oak Ridge, TN 37830
Phone: (877) 462-4873
Fax: (865) 220-1346



This Certificate will be accompanied by Calibration Charts or Readings where applicable

CUSTOMER INFORMATION				INSTRUMENT INFORMATION				
Customer Name: EnergySolutions Services, Inc.				Manufacturer: Ludlum				
Address: 1570 Bear Creek Road Oak Ridge, TN 37830				Model: 2360	Serial Number: 275724			
Contact Name: Tony Riggs				Probe: N/A	Serial Number: N/A			
Customer Purchase Order Number: N/A	Work Order Number: 2011-11716			Calibration Method: Electronic				
INSTRUMENT CALIBRATION INFORMATION								
Instrument Range	Calibration Standard Value	Ratemeter Response ($\pm 10\%$ of Standard Values)		Calibration Standard Value CPM	Time Base (minutes)	Tolerances (cpm) $\pm 2\%$	Scaler Response	
		As Found	As Left				As Found	As Left
X 1	100	100	100	1,000 CPM	0.1	90 – 110	99	99
X 1	250	250	250	1,000 CPM	0.5	450 – 550	493	493
X 1	400	400	400	1,000 CPM	1	900 – 1,100	988	988
X 10	1,000	1,000	1,000	1,000 CPM	2	1.8K–2.2K	1,979	1,979
X 10	2,500	2,500	2,500	1,000 CPM	5	4.5K-5.5K	4,948	4,948
X 10	4,000	4,000	4,000	1,000 CPM	10	9K-11K	9,896	9,896
X 100	10,000	10,000	10,000					
X 100	25,000	25,000	25,000					
X 100	40,000	40,000	40,000					
X 1000	100,000	100,000	100,000	Calibrated in accordance with OEM Technical Manual				
X 1000	250,000	250,000	250,000					
X 1000	400,000	400,000	400,000					
STATEMENT OF CERTIFICATION								
We Certify that the instrument listed above was calibrated and inspected prior to shipment and that it met all the Manufacturers published operating specifications. We further certify that our Calibration Measurements are traceable to the National Institute of Standards and Technology. (We are not responsible for damage incurred during shipment or use of this instrument).								
Instrument				Reviewed By: 				
Calibrated By: 				Date: 10/6/11				
Calibration Date: 10/06/2011				*Calibration Due (6mo): 04/06/2012				
				*Calibration Due (12mo): 10/06/2012				

* Calibration due date is dependant on users regulatory requirements.

Model: 2360



Serial Number: 275724

M&TE				Environmental Conditions		
Volt Meter	ID# 94710023	Cal Due: 10/28/2011		Barometer	ID# 2551	Cal Due: 10/19/2011
Pulser	ID# 120935	Cal Due: 09/29/2012		Thermometer	ID# 2551	Cal Due: 10/19/2011
Humidity	ID# 958670	Cal Due: 06/07/2012		Temp: 21.4 °C	Pressure: 747 mmHg	Humidity: 45%
Special Test						
BAT Check		Sat (√) Unsat ()		Geotropism		Sat (√) Unsat ()
LCD Display Check		Sat (√) Unsat ()		Audio Check		Sat (√) Unsat ()
Mechanical Zero		Sat (√) Unsat ()		Low BAT Set		Sat (√) Unsat ()
Reset		Sat (√) Unsat ()				
HV Analog Display		Sat (√) Unsat ()		As Found		As Left
High Voltage Calibration (± 10%)				Alpha Sensitivity = 128 mv		Alpha Sensitivity = 120 mv
Voltage	Tolerance	As Found	As Left	Beta Sensitivity = 3.7 mv		Beta Sensitivity = 3.5 mv
500	450-550	502	502	Beta Window = 30.8 mv		Beta Window = 30 mv
1000	900-1100	1,018	1,018	Beta Setpoints--Pulser counts detected at 3.5mv ± 1mv and shut off at 30mv for beta. For Alpha channel counts detected at 120mv and above.		
1500	1350-1650	1,514	1,514			
H.V. Set With Detector Not Connected				Overload to be set with detector to be used		
COMMENTS						
<p>Calibrated in accordance with OEM Technical Manual</p> <p>See detector certificate for High Voltage setting</p> <p>**Calibrated with 5ft cable**</p>						
Instrument				Reviewed By: 		
Calibrated By: 				Date: 10/6/11		
Calibration Date: 10/06/2011				*Calibration Due (6mo): 04/06/2012		
				*Calibration Due (12mo): 10/06/2012		

* Calibration due date is dependant on users regulatory requirements.

EnergySolutions Instrument Services
1570 Bear Creek Road
Oak Ridge, TN 37830
Phone: (877) 462-4873
Email: ISFstaff@energysolutions.com



This Certificate will be accompanied by Calibration Charts or Readings where applicable

CUSTOMER INFORMATION				DETECTOR INFORMATION				
Customer Name: EnergySolutions Instrument Services				Manufacturer: Ludlum				
Address: 1570 Bear Creek Road Oak Ridge, TN 37830				Detector Model: 43-37B				
Contact Name: Tony Riggs				Serial Number: 092501				
Customer Purchase Order Number: N/A		Work Order Number: 2012-12264		Evaluation Method: Source				
DETECTOR EFFICIENCY/RESPONSE/PRECISION INFORMATION								
Source Nuclide: C ¹⁴		Serial Number: 010002		Activity: 260,460 dpm		2 Pi Emissions: N/A	Certification Date:	
Parameter	As Found	As Left	Precision Test			CPM		
Count 1	32,658	32,658	Count 1 (Heel)			32,658		
Count 2	32,097	32,097	Count 2 (Center)			32,041		
Count 3	32,006	32,006	Count 3 (Toe)			33,092		
Count 4	32,041	32,041	Average			32,597		
Count 5	32,759	32,759	Tolerance			±10%		
Count 6	33,092	33,092	Pass/Fail			Pass		
Average	32,442.2	32,442.2						
Background (CPM)	503	503						
Net Counts	31,939.2	31,939.2						
2pi Efficiency	N/A	N/A						
4pi Efficiency	12.3%	12.3%						
Low Sample Activity: Source #: N/A		High Sample Activity: Source #: N/A		Dead Time (DT): N/A		Calibration Constant (CC): N/A		
SCALER INFORMATION			DETECTOR INFORMATION					
Model	Serial Number	Due Date	Background (cpm)	Operating Voltage		Threshold		
2360	275724	10/06/2012	503	1675V		Alpha (120mV) Beta (3.5-30mV)		
Detector Setup Report		YES NO ✓	Barcode Report		YES NO ✓	Voltage Plateau YES ✓ NO		
COMMENTS								
5 minute background performed Efficiency performed on contact with 5Ft. cable 1 layer mylar (0.4mg/cm2)								
STATEMENT OF CERTIFICATION								
We Certify that the detector listed above was evaluated for proper operation prior to shipment and that it met all the Manufacturers published operating specifications. We further certify that our Calibration Measurements are traceable to the National Institute of Standards and Technology. (We are not responsible for damage incurred during shipment or use of this detector).								
Detector		Certified By: 			Reviewed By: 			Date: 5/30/12
Certification Date: 05/30/2012				*Certification Due (6mo): 11/30/2012				
				* Certification Due (12mo): 05/30/2013				

* Calibration due date is dependant on users regulatory requirements.

EnergySolutions Instrument Services
1570 Bear Creek Road
Oak Ridge, TN 37830
Phone: (877) 462-4873
Email: ISFstaff@energysolutions.com

This Certificate will be accompanied by Calibration Charts or Readings where applicable

CUSTOMER INFORMATION				DETECTOR INFORMATION			
Customer Name: EnergySolutions Instrument Services				Manufacturer: Ludlum			
Address: 1570 Bear Creek Road Oak Ridge, TN 37830				Detector Model: 43-37A			
Contact Name: Tony Riggs				Serial Number: 092501			
Customer Purchase Order Number: N/A		Work Order Number: 2012-12264		Evaluation Method: Source			
DETECTOR EFFICIENCY/RESPONSE/PRECISION INFORMATION							
Source Nuclide: Th ²³⁰	Serial Number: 119738		Activity: 18,600 dpm		2 Pi Emissions: 8,640/min		Certification Date: 10/20/97
Parameter	As Found	As Left	Precision Test			CPM	
Count 1	2,899	2,899	Count 1 (Heel)			2,899	
Count 2	2,927	2,927	Count 2 (Center)			2,811	
Count 3	2,811	2,811	Count 3 (Toe)			2,968	
Count 4	3,022	3,022	Average			2,892.7	
Count 5	2,968	2,968	Tolerance			±10%	
Count 6	2,864	2,864	Pass/Fail			Pass	
Average	2,915.2	2,915.2					
Background (CPM)	6.9	6.9					
Net Counts	2,908.3	2,908.3					
2pi Efficiency	33.7%	33.7%					
4pi Efficiency	15.6%	15.6%					
Low Sample Activity: Source #: N/A		High Sample Activity: Source #: N/A		Dead Time (DT): N/A		Calibration Constant (CC): N/A	
SCALER INFORMATION			DETECTOR INFORMATION				
Model	Serial Number	Due Date	Background (cpm)	Operating Voltage		Threshold	
2360	275724	10/06/2012	6.9	1675V		Alpha (120mV) Beta (3.5-30mV)	
Detector Setup Report		YES NO ✓	Barcode Report		YES NO ✓	Voltage Plateau YES ✓ NO	
COMMENTS							
5 minute background performed Efficiency performed on contact with 5Ft. cable 1 layer mylar (0.4mg/cm2)							
STATEMENT OF CERTIFICATION							
We Certify that the detector listed above was evaluated for proper operation prior to shipment and that it met all the Manufacturers published operating specifications. We further certify that our Calibration Measurements are traceable to the National Institute of Standards and Technology. (We are not responsible for damage incurred during shipment or use of this detector).							
Detector							
Certified By: 		Reviewed By: 		Date: 5/30/12			
Certification Date: 05/30/2012				*Certification Due (6mo): 11/30/2012			
				* Certification Due (12mo): 05/30/2013			

* Calibration due date is dependant on users regulatory requirements.



**CALIBRATION
CERTIFICATE**

EnergySolutions Instrument Services
 1570 Bear Creek Road
 Oak Ridge, TN 37830
 Phone: (877) 462-4873
 Email: ISFstaff@energysolutions.com

This Certificate will be accompanied by Calibration Charts or Readings where applicable

CUSTOMER INFORMATION				DETECTOR INFORMATION					
Customer Name: EnergySolutions Instrument Services				Manufacturer: Ludlum					
Address: 1570 Bear Creek Road Oak Ridge, TN 37830				Detector Model: 43-37B					
Contact Name: Tony Riggs				Serial Number: 092501					
Customer Purchase Order Number: N/A		Work Order Number: 2012-12264		Evaluation Method: Source					
DETECTOR EFFICIENCY/RESPONSE/PRECISION INFORMATION									
Source Nuclide: Tc ⁹⁹		Serial Number: 099608		Activity: 21,312 dpm		2 Pi Emissions: 10,500/min	Certification Date: 08/08/96		
Parameter	As Found	As Left	Precision Test			CPM			
Count 1	4,820	4,820	Count 1 (Heel)			4,820			
Count 2	4,983	4,983	Count 2 (Center)			4,922			
Count 3	4,922	4,922	Count 3 (Toe)			5,003			
Count 4	4,767	4,767	Average			4,915			
Count 5	4,857	4,857	Tolerance			±10%			
Count 6	5,003	5,003	Pass/Fail			Pass			
Average	4,892	4,892							
Background (CPM)	503	503							
Net Counts	4,389	4,389							
2pi Efficiency	41.8%	41.8%							
4pi Efficiency	20.6%	20.6%							
Low Sample Activity: Source #: N/A		High Sample Activity: Source #: N/A		Dead Time (DT): N/A		Calibration Constant (CC): N/A			
SCALER INFORMATION			DETECTOR INFORMATION						
Model	Serial Number	Due Date	Background (cpm)	Operating Voltage		Threshold			
2360	275724	10/06/2012	503	1675V		Alpha (120mV) Beta (3.5-30mV)			
Detector Setup Report		YES	NO	✓	Barcode Report		YES	NO	✓
COMMENTS									
5 minute background performed		Efficiency performed on contact with 5Ft. cable			1 layer mylar (0.4mg/cm2)				
STATEMENT OF CERTIFICATION									
We Certify that the detector listed above was evaluated for proper operation prior to shipment and that it met all the Manufacturers published operating specifications. We further certify that our Calibration Measurements are traceable to the National Institute of Standards and Technology. (We are not responsible for damage incurred during shipment or use of this detector).									
Detector									
Certified By:		Reviewed By:		Date: 5/30/12					
Certification Date: 05/30/2012				*Certification Due (6mo): 11/30/2012					
				* Certification Due (12mo): 05/30/2013					

* Calibration due date is dependant on users regulatory requirements.



Designer and Manufacturer
of
Scientific and Industrial
Instruments

CERTIFICATE OF CALIBRATION

LUDLUM MEASUREMENTS, INC.
POST OFFICE BOX 810 PH. 325-235-5494
501 OAK STREET FAX NO. 325-235-4672
SWEETWATER, TEXAS 79556, U.S.A.

CUSTOMER US ARMY CORPS OF ENG BALTIMORE ORDER NO. 20181694/366617

Mfg. Ludlum Measurements, Inc. Model 19 Serial No. 245026

Mfg. _____ Model _____ Serial No. _____

Cal. Date 11-Aug-11 Cal Due Date 11-Aug-12 Cal. Interval 1 Year Meterface 202-1070

Check mark applies to applicable instr. and/or detector IAW mfg. spec. T. 74 °F RH 41 % Alt 698.8 mm Hg

New Instrument Instrument Received Within Toler. +10% 10-20% Out of Tol. Requiring Repair Other-See comments

Mechanical ck. Meter Zeroed Background Subtract Input Sens. Linearity

F/S Resp. ck. Reset ck. Window Operation Geotropism

Audio ck. Alarm Setting ck. Batt. ck. (Min. Volt) 2.2 VDC

Calibrated in accordance with LMI SOP 14.8 rev 12/05/89. Calibrated in accordance with LMI SOP 14.9 rev 02/07/97.

Instrument Volt Set 650 V Input Sens. 30 mV Det. Oper. _____ V at _____ mV Threshold Dial Ratio _____ = _____ mV

HV Readout (2 points) Ref./Inst. _____ / _____ V Ref./Inst. _____ / _____ V

COMMENTS:

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING*
5000	4000µR/hr	N/A	4000
5000	1000µR/hr		1000
500	400µR/hr = 76,300 cpm		400
500	100µR/hr		100
250	200µR/hr = 38,000 cpm		200
250	100µR/hr		100
50	7680 cpm		40
50	1920 cpm		10
25	3800 cpm		20
25	950 cpm		5

Range(s) Calibrated Electronically

*Uncertainty within ± 10% C.F. within ± 20%

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*
Digital Readout			Log Scale		

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements of ANSI/NCSL Z540-1-1994 and ANSI N323-1978 State of Texas Calibration License No. LO-1963

Reference Instruments and/or Sources: 73410 1131 781 059 280 60646 70897
 Cs-137 Gamma S/N 1162 G112 M565 5105 T1008 T879 E552 E551 720 734 1616 Neutron Am-241 Be S/N T-304
 Alpha S/N _____ Beta S/N _____ Other _____
 m 500 S/N 94940 Oscilloscope S/N _____ Multimeter S/N 78401031

Calibrated By: [Signature] Date 11-Aug-11

Reviewed By: [Signature] Date 11 Aug 11

AC Inst. Passed Dielectric (Hi-Pot) and Continuity Test
Only Failed: _____



Designer and Manufacturer
of
Scientific and Industrial
Instruments

CERTIFICATE OF CALIBRATION

LUDLUM MEASUREMENTS, INC.
POST OFFICE BOX 810 PH. 325-235-5494
501 OAK STREET FAX NO. 325-235-4672
SWEETWATER, TEXAS 79556, U.S.A.
20181694/366617
~~20181694/366614~~

CUSTOMER US ARMY CORPS OF ENG BALTIMORE ORDER NO. _____
Mfg. Ludlum Measurements, Inc. Model 19 Serial No. 253050
Mfg. _____ Model _____ Serial No. _____
Cal. Date 11-Aug-11 Cal Due Date 11-Aug-12 Cal. Interval 1 Year Meterface 202-1070

Check mark applies to applicable instr. and/or detector IAW mfg. spec. T. 74 °F RH 38 % Alt 698.8 mm Hg
 New Instrument Instrument Received Within Toler. +10% 10-20% Out of Tol. Requiring Repair Other-See comments
 Mechanical ck. Meter Zeroed Background Subtract Input Sens. Linearity
 F/S Resp. ck. Reset ck. Window Operation Geotropism
 Audio ck. Alarm Setting ck. Batt. ck. (Min. Volt) 2.2 VDC
 Calibrated in accordance with LMI SOP 14.8 rev 12/05/89. Calibrated in accordance with LMI SOP 14.9 rev 02/07/97.
Instrument Volt Set 686 V Input Sens. 29 mV Det. Oper. _____ V at _____ mV Threshold Dial Ratio _____ = _____ mV
 HV Readout (2 points) Ref./Inst. _____ / _____ V Ref./Inst. _____ / _____ V

COMMENTS:

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING*
5000	4000µR/hr	4100	4000
5000	1000µR/hr	1100	1000
500	400µR/hr = 76,000 cpm	395	400
500	100µR/hr	95	100
250	200µR/hr = 37,600 cpm	210	200
250	100µR/hr	110	100
50	7600 cpm	39.5	40
50	1900 cpm	9.5	10
25	3760 cpm	19.5	20
25	940 cpm	4.5	5

Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements of ANSI/NCSL Z540-1-1994 and ANSI N323-1978 State of Texas Calibration License No. LO-1963

Reference Instruments and/or Sources: 73410 1131 781 059 280 60646 70897
Cs-137 Gamma S/N 1162 G112 M565 5105 T1008 T879 E552 E551 720 734 1616 Neutron Am-241 Be S/N T-304
 Alpha S/N _____ Beta S/N _____ Other _____
 m 500 S/N 94940 Oscilloscope S/N _____ Multimeter S/N 78401031

Calibrated By: Scott S. Brown Date 11-Aug-11
Reviewed By: Rhonda H. H. Date 11 Aug 11

AC Inst. Only	<input type="checkbox"/> Passed Dielectric (Hi-Pot) and Continuity Test
	<input type="checkbox"/> Failed:



EBERLINE SERVICES

CERTIFICATE OF CALIBRATION

Electroplated Alpha Standard

S.O.# 7155
P.O.# C/C

Description of Standard:

Model No. DNS-11 Serial No. 7241-11 Isotope Th-230

Electroplated on polished SS disc, 0.79 mm thick.

Total diameter of 4.77 cm and an active diameter of 4.45 cm.

The radioactive material is permanently fixed to the disc by heat treatment without any covering over the active surface.

Measurement Method:

The 2pi alpha emission rate was measured using an internal gas flow proportional chamber. Absolute counting of alpha particles emitted in the hemisphere above the active surface was verified by counting above, below, and at the operative voltage. The calibration is traceable to NIST by reference to an NIST calibrated alpha source S/N 75322-201

Measurement Result:

The observed alpha particles emitted from the surface of the disc per minute (cpm) on the calibration date was:

11,300 + 338

The total disintegration rate (dpm) assuming 1.5% backscatter of alpha particles from the surface of the disc, was:

22,200 + 666 (0.0100 μ Ci)

The uncertainty of the measurement is 3 %, which is the sum of random counting error at the 99% confidence level, and the estimated upper limit of systematic error in this measurement.

Calibrated by: ART REUST

Reviewed by: [Signature]

Calibration Technician: [Signature]

Q.A. Manager: [Signature]

Calibration Date: 10-03-2011

Reviewed Date: 10/10/2011

Source Manufacturing Lab
7021 Pan American Freeway NE
Albuquerque, New Mexico 87109-4238
(505) 761-5413 Fax (505) 761-5416
art.reust@eberlineservices.com



EBERLINE SERVICES

CERTIFICATE OF CALIBRATION

Electroplated Alpha Standard

S.O.# 7155
P.O.# C/C

Description of Standard:

Model No. DNS-11 Serial No. 7242-11 Isotope Th-230

Electroplated on polished SS disc, 0.79 mm thick.

Total diameter of 4.77 cm and an active diameter of 4.45 cm.

The radioactive material is permanently fixed to the disc by heat treatment without any covering over the active surface.

Measurement Method:

The 2pi alpha emission rate was measured using an internal gas flow proportional chamber. Absolute counting of alpha particles emitted in the hemisphere above the active surface was verified by counting above, below, and at the operative voltage. The calibration is traceable to NIST by reference to an NIST calibrated alpha source S/N 75322-201

Measurement Result:

The observed alpha particles emitted from the surface of the disc per minute (cpm) on the calibration date was:

14,600 ± 438

The total disintegration rate (dpm) assuming 1.5% backscatter of alpha particles from the surface of the disc, was:

28,800 ± 863 (0.0130 μCi)

The uncertainty of the measurement is 3%, which is the sum of random counting error at the 99% confidence level, and the estimated upper limit of systematic error in this measurement.

Calibrated by: ART REUST

Reviewed by: [Signature]

Calibration Technician: [Signature]

Q.A. Manager: [Signature]

Calibration Date: 10-03-2011

Reviewed Date: 10/10/2011

Source Manufacturing Lab
7021 Pan American Freeway NE
Albuquerque, New Mexico 87109-4238
(505) 761-5413 Fax (505) 761-5416
art.reust@eberlineservices.com



EBERLINE SERVICES

CERTIFICATE OF CALIBRATION

Electroplated Alpha Standard

S.O.# 7155
P.O.# C/C

Description of Standard:

Model No. DNS-11 Serial No. 7244-11 Isotope Th-230

Electroplated on polished SS disc, 0.79 mm thick.

Total diameter of 4.77 cm and an active diameter of 4.45 cm.

The radioactive material is permanently fixed to the disc by heat treatment without any covering over the active surface.

Measurement Method:

The 2pi alpha emission rate was measured using an internal gas flow proportional chamber. Absolute counting of alpha particles emitted in the hemisphere above the active surface was verified by counting above, below, and at the operative voltage. The calibration is traceable to NIST by reference to an NIST calibrated alpha source S/N 75322-201

Measurement Result:

The observed alpha particles emitted from the surface of the disc per minute (cpm) on the calibration date was:

12,200 ± 365

The total disintegration rate (dpm) assuming 1.5% backscatter of alpha particles from the surface of the disc, was:

24,000 ± 720 (0.0108 μCi)

The uncertainty of the measurement is 3%, which is the sum of random counting error at the 99% confidence level, and the estimated upper limit of systematic error in this measurement.

Calibrated by: ART REUST

Reviewed by: [Signature]

Calibration Technician: [Signature]

Q.A. Manager: [Signature]

Calibration Date: 10-03-2011

Reviewed Date: 10/10/2011

Source Manufacturing Lab
7021 Pan American Freeway NE
Albuquerque, New Mexico 87109-4238
(505) 761-5413 Fax (505) 761-5416
art.reust@eberlineservices.com



EBERLINE SERVICES

CERTIFICATE OF CALIBRATION

Electroplated Beta Standard

S.O.# 7156
P.O.# C/C

Description of Standard:

Model No. DNS-12 Serial No. 7246-11 Isotope Tc-99

Electroplated on polished SS disc, 0.79 mm thick.

Total diameter of 4.77 cm and an active diameter of 4.45 cm.

The radioactive material is permanently fixed to the disc by heat treatment without any covering over the active surface.

Measurement Method:

The 2pi beta emission rate was measured using an internal gas flow proportional chamber. Absolute counting of beta particles emitted in the hemisphere above the active surface was verified by counting above, below, and at the operative voltage. The calibration is traceable to NIST by reference to an NIST calibrated beta source S/N 75323-201.

Measurement Result:

The observed beta count rate from the surface of the disc per minute (cpm) on the calibration date was:

11,500 + 458

The total disintegration rate (dpm) assuming 25 % backscatter of beta particles from the surface of the disc, was:

18,300 + 733 (0.00826 μ Ci)

The uncertainty of the measurement is 4 %, which is the sum of random counting error at the 99% confidence level, and the estimated upper limit of systematic error in this measurement.

Calibrated by: ART REUST

Reviewed by: [Signature]

Calibration Technician: [Signature]

Q.A. Manager: [Signature]

Calibration Date: 10-03-2011

Reviewed Date: 10/10/2011

Source Manufacturing Lab
7021 Pan American Freeway NE
Albuquerque, New Mexico 87109-4238
(505) 761-5413 Fax (505) 761-5416
art.reust@eberlineservices.com



EBERLINE SERVICES

CERTIFICATE OF CALIBRATION

Electroplated Beta Standard

S.O.# 7156
P.O.# C/C

Description of Standard:

Model No. DNS-12 Serial No. 7245-11 Isotope Tc-99

Electroplated on polished SS disc, 0.79 mm thick.

Total diameter of 4.77 cm and an active diameter of 4.45 cm.

The radioactive material is permanently fixed to the disc by heat treatment without any covering over the active surface.

Measurement Method:

The 2pi beta emission rate was measured using an internal gas flow proportional chamber. Absolute counting of beta particles emitted in the hemisphere above the active surface was verified by counting above, below, and at the operative voltage. The calibration is traceable to NIST by reference to an NIST calibrated beta source S/N 75323-201.

Measurement Result:

The observed beta count rate from the surface of the disc per minute (cpm) on the calibration date was:

6,600 + 329

The total disintegration rate (dpm) assuming 25 % backscatter of beta particles from the surface of the disc, was:

10,600 + 527 (0.00475 μ Ci)

The uncertainty of the measurement is 5 %, which is the sum of random counting error at the 99% confidence level, and the estimated upper limit of systematic error in this measurement.

Calibrated by: ART REUST

Reviewed by: [Signature]

Calibration Technician: [Signature]

Q.A. Manager: [Signature]

Calibration Date: 10-03-2011

Reviewed Date: 10/10/2011

Source Manufacturing Lab
7021 Pan American Freeway NE
Albuquerque, New Mexico 87109-4238
(505) 761-5413 Fax (505) 761-5416
art.reust@eberlineservices.com



EBERLINE SERVICES

CERTIFICATE OF CALIBRATION

Electroplated Beta Standard

S.O.# 7156
P.O.# C/C

Description of Standard:

Model No. DNS-12 Serial No. 7248-11 Isotope Tc-99

Electroplated on polished SS disc, 0.79 mm thick.

Total diameter of 4.77 cm and an active diameter of 4.45 cm.

The radioactive material is permanently fixed to the disc by heat treatment without any covering over the active surface.

Measurement Method:

The 2pi beta emission rate was measured using an internal gas flow proportional chamber. Absolute counting of beta particles emitted in the hemisphere above the active surface was verified by counting above, below, and at the operative voltage. The calibration is traceable to NIST by reference to an NIST calibrated beta source S/N 75323-201.

Measurement Result:

The observed beta count rate from the surface of the disc per minute (cpm) on the calibration date was:

10,800 + 430

The total disintegration rate (dpm) assuming 25 % backscatter of beta particles from the surface of the disc, was:

17,200 + 688 (0.00775 μ Ci)

The uncertainty of the measurement is 4 %, which is the sum of random counting error at the 99% confidence level, and the estimated upper limit of systematic error in this measurement.

Calibrated by: ART REUST Reviewed by: [Signature]

Calibration Technician: [Signature] Q.A. Manager: [Signature]

Calibration Date: 10-03-2011 Reviewed Date: 10/10/2011

Source Manufacturing Lab
7021 Pan American Freeway NE
Albuquerque, New Mexico 87109-4238
(505) 761-5413 Fax (505) 761-5416
art.reust@eberlineservices.com