

Simmons, Michelle

From: Simmons, Michelle
Sent: Wednesday, October 24, 2012 4:17 PM
To: 'aiaroseann@acsalaska.net'
Subject: PET/CT Center of Alaska Deficiency Letter mail Control No. 579108

Docket: 030-38575
License: 50-35006-01
Control: 579108

This is in reference to your application received September 4, 2012, requesting a new Nuclear Regulatory Commission License No. 50-35006-01. In order to continue our review, we need the following additional information:

Please commit the following statements:

- "Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations."
- "We have developed and will implement and maintain written survey meter calibration procedures in accordance with the requirements in 10 CFR 20.1501 and that meet the requirements of 10 CFR 35.61."
- "Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions."
- "Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under "Criteria" in NUREG- 1556, Vol. 9, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licensees," dated October 2002."
- "We have developed and will implement and maintain procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1301."

Please submit the additional information by October 31, 2012. If you are unable to respond by this due date, please don't hesitate to contact me so we can discuss an extension to the date. Our fax number is (817) 200-1263. You may also email your response as a pdf attachment. If you have any questions regarding this email, please call me at 817-200-1590. When responding to this email, please include the license, docket and control numbers located at the top of this page.

Thanking you in advance for your cooperation, assistance, and prompt response in this matter

Michelle Simmons
Health Physicist
U.S. Nuclear Regulatory Commission
Division of Nuclear Materials Safety
Region IV
1600 East Lamar Boulevard
Arlington, TX 76011-4511
817-200-1590

PUBLIC
 Immediate Release
 Normal Release

NON-PUBLIC
 A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other: _____

Reviewer: MS Date: 10/24/12

ML12298A565