

<p>NRC FORM 314 (12-2010) 10 CFR 30.36(j)(1); 40 CFR 61.11; 70 CFR 101.11 and 72 CFR 101.11</p>	<p>U.S. NUCLEAR REGULATORY COMMISSION</p>	<p>APPROVED BY OMB: NO. 3150-0028 EXPIRES: 10/31/2013</p> <p><small>Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Information Services Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to Infocollects.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NECB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.</small></p>
--	--	--

CERTIFICATE OF DISPOSITION OF MATERIALS

<p>LICENSEE NAME AND ADDRESS Paseo Corridor Constructors 103 West 26th Avenue North Kansas City, Missouri</p>	<table style="width:100%;"> <tr> <td style="width:50%;">LICENSE NUMBER 24-32689-01</td> <td style="width:50%;">DOCKET NUMBER 03037661</td> </tr> <tr> <td colspan="2">LICENSE EXPIRATION DATE 10/02/2011</td> </tr> </table>	LICENSE NUMBER 24-32689-01	DOCKET NUMBER 03037661	LICENSE EXPIRATION DATE 10/02/2011	
LICENSE NUMBER 24-32689-01	DOCKET NUMBER 03037661				
LICENSE EXPIRATION DATE 10/02/2011					

A. LICENSE STATUS (Check the appropriate box)

This license has expired. This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

1. No radioactive materials have ever been procured or possessed by the licensee under this license.
2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner:
 - a. Transfer of radioactive materials to the licensee listed below:
KDHE License No. 22-B845
 - b. Disposal of radioactive materials:
 - 1. Directly by the licensee:
Date of transfer: August 5, 2011
Geotechnology, Inc., Steve Birtz 913-438-1900
 - 2. By licensed disposal site:
 - 3. By waste contractor:
- c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

1. A radiation survey was conducted by the licensee. The survey confirms:
 - a. the absence of licensed radioactive materials
 - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
2. A copy of the radiation survey results:
 - a. is attached; or
 - b. is not attached (Provide explanation); or
 - c. was forwarded to NRC on: _____ Date
3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
 - a. The results of the latest leak test are attached; and/or
 - b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME Jontell Jones	TITLE RSO	TELEPHONE (Include Area Code) (816)500-2864	E-MAIL ADDRESS jjones@clarksonconstruction.com
-----------------------	--------------	--	---

Mail all future correspondence regarding this license to:
4133 Gardner Avenue, Kansas City, MO 64120

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE Jontell Jones Quality Control Project Manager	SIGNATURE 	DATE 10/29/2012
---	---------------	--------------------

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

Form: 193A Rev 1, 19 Jun 2006



www.qaltek.com
Qal-Tek Associates

7801 N. Lamar Blvd. Ste. E204
Austin, TX 78752
Ph: 512 407-9252

3998 Commerce Circle
Idaho Falls, ID 83401
Ph: 208 523-5557
Fax: 208 524-8470

712 S. Hacienda Drive Ste. 6
Tempe, AZ 85281
Ph: 480 304-5199

SEALED RADIOACTIVE SOURCE LEAK TEST REPORT

Company: Paseo Corridor Construction Acct#: 102377 Ref#: W288-001180
Street: 103 W. 26th Ave
City/ST/Zip: N. Kansas City, MO 64116
Phone: 816-423-8200 Fax: 816-423-8202
Serial Number: **38733** Mfg: Troxler Model: 3440
Leak Test Date: 21-Jul-2011 Leak Test Due Date: 21-Jan-2012

Test Instrument

Test Instrument	Ludlum : 2360 ; 220261
-----------------	------------------------

Test Instrument Info

MDA	less than 0.005 µCi
α efficiency	41
β efficiency	32
Det. Type	ZnS (Ag)
Last Cal. Date	10-Aug-2010

Test Results

	Isotope	Activity	net a CPM	net b/g CPM	net a µCi	net b/g µCi	Pass/Fail
	Am241	40 mCi	0	0	0	0	Pass
	Cs137	8 mCi	0	1	0	0.000003	Pass

Instrument Technician: Travis Bivins
Date of Service: 21-Jul-2011
Next Leak Test Due: 21-Jan-2012

Qal-Tek Associates certifies that all leak test measurements are performed in accordance with NRC licensee requirements for isotopic detection limits. For this purpose the MDA is below the NRC regulatory limits of <0.005 µCi

Form: 193A Rev 1, 19 Jun 2006



www.qaltek.com
Qal-Tek Associates

7801 N. Lamar Blvd. Ste. E204
Austin, TX 78752
Ph: 512 407-9252

3998 Commerce Circle
Idaho Falls, ID 83401
Ph: 208 523-5557
Fax: 208 524-8470

712 S. Hacienda Drive Ste. 6
Tempe, AZ 85281
Ph: 480 304-5199

SEALED RADIOACTIVE SOURCE LEAK TEST REPORT

Company: Paseo Corridor Construction Acct#: 102377 Ref#: W288-001182
Street: 103 W. 26th Ave
City/State/Zip: N. Kansas City, MO 64116
Phone: 816-423-8200 Fax: 816-423-8202
Serial Number: 28413 Mig: Troxler Model: 3430
Leak Test Date: 21-Jul-2011 Leak Test Due Date: 21-Jan-2012

Test Instrument

Test Instrument	Ludlum : 2360 : 220261
-----------------	------------------------

Test Instrument Info

MDA	less than 0.005 µCi
α efficiency	41
β efficiency	32
Det. Type	ZnS (Ag)
Last Cal. Date	10-Aug-2010

Test Results

	Isotope	Activity	net a CPM	net b/g CPM	net a µCi	net b/g µCi	Pass/Fail
	Am241	40 mCi	0	0	0	0	Pass
	Cs137	8 mCi	1	0	0.000002	0	Pass

Instrument Technician: Travis Bivins
Date of Service: 21-Jul-2011
Next Leak Test Due: 21-Jan-2012

Qal-Tek Associates certifies that all leak test measurements are performed in accordance with NRC licensee requirements for isotopic detection limits. For this purpose the MDA is below the NRC regulatory limits of <0.005 µCi

Form: 193A Rev 1, 19 Jun 2005



www.qaltek.com
Qal-Tek Associates

7801 N. Lamar Blvd. Ste. B204
Austin, TX 78752
Ph: 512 407-9252

3998 Commerce Circle
Idaho Falls, ID 83401
Ph: 208 523-5557
Fax: 208 524-8470

712 S. Hacienda Drive Ste. 6
Tempe, AZ 85281
Ph: 480 304-5199

SEALED RADIOACTIVE SOURCE LEAK TEST REPORT

Company: Paseo Corridor Construction Acct#: 102377 Ref#: W288-001183
Street: 103 W. 26th Ave
City/St/Zip: N. Kansas City, MO 64116
Phone: 816-423-8200 Fax: 816-423-8202
Serial Number: **27686** Mfg: Troxler Model: 3440
Leak Test Date: 21-Jul-2011 Leak Test Due Date: 21-Jan-2012

Test Instrument

Test Instrument	Ludlum : 2360 : 220261
-----------------	------------------------

Test Instrument Info

MDA	less than 0.005 µCi
α efficiency	41
β efficiency	32
Det. Type	ZnS (Ag)
Last Cal. Date	10-Aug-2010

Test Results

Isotope	Activity	net a CPM	net b/g CPM	net a µCi	net b/g µCi	Pass/Fail
Am241	40 mCi	0	0	0	0	Pass
Cs137	8 mCi	0	0	0	0	Pass

Instrument Technician: Travis Bivins
Date of Service: 21-Jul-2011
Next Leak Test Due: 21-Jan-2012

Qal-Tek Associates certifies that all leak test measurements are performed in accordance with NRC licensee requirements for isotopic detection limits. For this purpose the MDA is below the NRC regulatory limits of <0.005 µCi



103 West 26th Avenue
 North Kansas City, MO 64116
 (816) 423-8200 (phone)
 (816) 423-8202 (fax)

August 5, 2011

Bill of Sale

This a notice of sale for Troxler Nuclear Gauges have been sold to Geotechnology, Inc. from Paseo Corridor Constructors.

<u>Gauge Model Number</u>	<u>Gauge Serial Number</u>
3440	38733
3440	27542
3430	28413
3440	27686

Radiation Safety Officer, PCC
 John Truesdale
 24-32689-01

Radiation Safety Officer, Geotechnolgy
 Steve Birtz
 KANE License No. 22-8845

Project Manager, PCC
 Bryan Wilkerson

Certificate of Completion

This certifies that

Gontell Jones

has successfully completed the
Radiation Safety Officer Training Class
conducted by the training department of

Troxler Electronic Laboratories, Inc.

P. Patrick Schroeder

Patrick Schroeder
Instructor

April 26, 2012

Date

William F. Troxler, Jr.
President



Troxler Electronic Laboratories, Inc.
PO Box 12057 * 3008 Cornwallis Rd. * Research Triangle Park NC 27709
Phone: (919) 548-8661 * Fax: (919) 548-0761 * Web site: www.troxlerlabs.com

41845895

HAZMAT Certification

as required by U.S. DOT and IATA

This certifies that
Jontell Jones

has been trained and tested in accordance with the U.S. Department of Transportation and International Air Transport Association (IATA) hazardous material requirements for general awareness/familiarization, function-specific, safety, and security awareness training as related to the transportation of nuclear gauges. A description of the training course materials is available from Troxler Electronic Laboratories, Inc.

Training Date
April 26, 2012

Expiration Date
3 years from date of class

Instructor
Patrick Schroeder



Troxler Electronic Laboratories, Inc.
PO Box 12057 • 3006 Cornwallis Road • Research Triangle Park, NC 27708
Phone: (919) 549-8881 • Fax: (919) 549-0781 • www.troxlerlabs.com

Hazmat Employer Certification

Company:

Company Official: Manu K. Rupard Date: 5/2/12

Enrollment ID: 41845895

Clarkson Construction Company
4133 Gardner Ave.
Kansas City, Missouri 64120

(816)483-8800
Fax 816-241-6023
jjones@clarksonconstruction.com

TO:	Frank Tran	FROM:	Jontell Jones
FAX:	630-515-1259	PAGES:	9
PHONE:	630-829-9623	DATE:	October 29, 2012
RE:	License 24-32689-01 Paseo Corridor Constructors	CC:	Michael Herr

URGENT FOR REVIEW PLEASE COMMENT PLEASE REPLY PLEASE RECYCLE

Attached form 314; CERTIFICATE OF DISPOSITION OF MATERIALS, Leak Test Reports, RSO Certificate.

Please let me know if you have any questions or concerns,

Thanks

Jontell Jones
Quality Control Project Manager, RSO
816-500-2864