



202 Maplewood Avenue · Ronceverte, West Virginia 24970
Phone (304) 647-4411 · Facsimile (304) 647-6010 · www.gvmc.com

10/23/12

U.S. Nuclear Regulatory Commission
Attn: Lester Tripp
Region I
2100 Renaissance BLVD
King of Prussia, PA 19406

L 9
030/2343

Licensee: Greenbrier Valley Medical Center
License No. 47-17199-01
Mail Control No. 577920

Mr. Tripp

This letter is written in response to your request for additional information on 10/17/12. All responses are in order of request.


1. NA
2. Two survey instruments are available for use. Current probe models used for all surveys including surveys for brachytherapy seeds are as follows:

Ludlum GM Model 3	Ludlum Model 44-9 Pancake G-M Detector
Ludlum GM Model 3	Ludlum Model 44-7 is a Thin End Window G-M

3. We confirm that patients administered radioactive materials identified in 10 CFR 35.300 and 35.400 will be releasable in accordance with 10 CFR 35.75.
4. Materials identified in 10 CFR 31.11 have never been used at this facility.
5. NA
6. Please remove Charles E. Gabe, D.O. from our license.

Thank you for your attention to this matter. If you have any questions, please contact Sharon L. Long, NPC at 888-456-5255 or slong@nationalphysics.com

Sincerely,


Administrator

577920
NMSS/RGN1 MATERIALS-002