

U.S. NUCLEAR REGULATORY COMMISSION

Date: 10/31/2012

TELEPHONE CONVERSATION RECORDMail Control
or Report No(s).

579038

License No(s).

SUB-1551

Docket No(s).

04009007

Name of Licensee:

Neutron Products, Inc.

Name of Participant(s):Edmond J. DeRosa, RSO
Dennis Lawyer, R1, DNMS**Telephone No.**

301-349-5001

Subject:

RAI, Instruments

(NOTE: This will be used as the
Documents Title in ADAMS)

I discussed with Mr. DeRosa that I needed the types of instruments being used at his facility. I did not see a statement in his calibration procedure that the sources used to calibrate was traceable to National Institute of Standards and Technology and that there was a efficiency calibration done on count rate instruments.

I also discussed that for question number 7 that I could not give authorization for his flexibility request as he was not a Broadscope A licensee, but he could re-answer the question with a confirmation that they have operating and emergency procedures.

Action Required: Mr. DeRosa will reply about the instrument types and calibrations and review question number 7.

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Sensitive – Privacy Act/ PII

Immediate Release

Non-Sensitive Copyright

Sensitive – Internal

Sensitive- Proprietary

Sensitive – Security-Related

Normal Release Date: 11/09/2012

Delayed Release Date:

SUNSI Review Completed By:

Dennis Lawyer/ RA /