

Sibley Memorial Hospital
5255 Loughboro Road, N.W.
Washington, D.C. 20016-2695
202-537-4000 T

J9



November 12, 2012

Licensing Assistant Section
Nuclear Materials Safety Branch
U.S. Nuclear Regulatory Commission
Region 1
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406

REC'D NOV 13 12 PM 08:04

RE: License Amendment Request
Sibley Memorial Hospital
NRC-08-07398-03 03014754

Dear License Reviewer:

We are requesting the release for unrestricted use of the areas previously utilized as the Brachytherapy Storage Room (storage location of the HDR Afterloader and brachytherapy sealed sources) and Vault 2 (location where HDR Afterloader cases were performed) at Sibley Memorial Hospital located at 5255 Loughboro Road NW, Washington DC 20016. In accordance with the guidelines established by the NRC, dated December 1975, "Guidelines for Decontamination of Facilities and Equipment Prior to Release for Unrestricted Use" and NUREG-1757 Vol. 1, Rev. 2 – Consolidated Decommissioning Guidance, we herein submit the final survey of these areas. This survey was performed by Ms. Michele Loscocco of Krueger-Gilbert Health Physics, Inc. on November 5, 2012. (Refer to Enclosure 1) Documentation of the most recent leak test is also enclosed. (Refer to Enclosure 2)

All radioactive material, including the HDR Afterloader and brachytherapy sealed sources were moved to the new Radiation Oncology Brachytherapy Storage Room located at the facility on November 5, 2012 by hospital staff. (Refer to Enclosure 3) The new storage room was added to the above referenced license per Amendment No. 38.

We request that a representative from the NRC perform the final closeout at the earliest possible convenience or that the areas be released based on the enclosed survey. Once the final closeout has been conducted and the areas released, it is also requested that these areas (Radiation Oncology Department spaces including Brachytherapy Storage Room and Vault 2) be removed from radioactive materials license NRC-08-07398-03. The building will be demolished upon receipt of final approval. Demolition is currently scheduled for December 1, 2012. Therefore, it is requested that the final release approval be expedited.

579469

License Amendment Request
Sibley Memorial Hospital
NRC-08-07398-03
Page 2 of 2

If you have any questions regarding this request, please contact the undersigned or Ms. Michele Loscocco of Krueger-Gilbert Health Physics, Inc. at (410) 692-9806.

Sincerely,



Jordie Keck, MS, DABR
Radiation Safety Officer

Enclosure 1 – Survey Results for Brachytherapy Storage Room
Enclosure 2 – Leak test results for HDR Afterloader Source
Enclosure 3 – Brachytherapy Source Inventory

Enclosure 1 - SURVEY RESULTS for Brachytherapy Storage Room

Facility: Sibley Memorial Hospital
Address: 5255 Loughboro Road, N.W.
Washington, D.C. 20016-2695

Survey date: November 5, 2012

Radiation exposure levels were monitored with a Ludlum 14C, #206513 survey meter last calibrated on September 5, 2012. Areas surveyed and results are indicated on the enclosed diagrams.

Wipe tests were conducted on 100 cm² areas using absorbent paper moistened with alcohol. Wipe testing and results are indicated on the enclosed diagrams.

Instrument: Ludlum Scaler Model 2200/Ludlum NaI Well Model 243

Detector Efficiency: Cs-137 (662 keV) = 12.0%
Ba-133 (356 keV) = 22.3%
Co-57 (122 keV) = 69.9%

The minimal detectable activity (MDA) was determined using a worst case efficiency of 12.0% and background + 3 σ .

MDA = 6.31×10^{-5} uCi or 118.5 dpm

All sample results were less than the minimal detectable activity.
Net Wipe test results were also less than 200 dpm/100 cm².

SURVEY RESULTS
Brachytherapy Storage Room

Survey Date: November 5, 2012
Instrument: Ludlum 14C, #206513 survey meter
Calibration Date: September 5, 2012
Background: 150 - 200 cpm


<u>Sample Location</u>	<u>Survey Meter Reading</u>
1 - 5	200 cpm
6 - 10	200 cpm
11 - 15	200 cpm
16 - 20	200 cpm

WIPE TEST RESULTS
Brachytherapy Storage Room

Survey Date: November 5, 2012
Instrument: Ludlum Scaler Model 2200/Ludlum NaI Well Model 243
Detector Efficiency: Cs-137 (662 keV) = 12.0%
Ba-133 (356 keV) = 49.7%
Co-57 (122 keV) = 96.0%

<u>Sample Location</u>	<u>Results (Net DPM)</u>
1 - 5	<MDA
6 - 10	<MDA
11 - 15	<MDA
16 - 20	<MDA

Survey locations marked on room diagram.



BRACHYTHERAPY STORAGE ROOM

Certificate For Sealed Sources

Issue Date: 2012-08-15 (1)

Product Code: REF 105002
Serial Number: SN D36E-2572
Production Code: LOT HDR080912

Serial no. Transport Container: 2303C6 ✓
Serial no. Check Cable: N/A
Certificate Number: gSvtM xy8@M o++Yk K03k2 s3

Source Specification

Reference Air Kerma Rate: 41.84 mGy h⁻¹ +/- 5% at 1m (2)
Measurement Reference Date: 2012-08-16 18:00 CET (1)
Apparent Activity: 380 GBq (10.3 Ci) at measurement reference date. (3,4)

Source Type: MICROSELECTRON-HDR
Capsule dimensions: 0.9 mm diame.er, 4.5 mm length
Source pellet dimensions: 0.6 mm diameter, 3.5 mm length
Source pellet form: Solid Iridium
Radionuclide: Ir-192
Encapsulation: single
Capsule material: stainless steel AISI 316L
ISO Classification: ISO/99/C6321 I
Special form certificate number: D/0070/S-96

Quality Control

Laser Weld Visual Check: Passed
Source Capsule Integrity (15 N pull test): Passed
Leakage Test: Passed (5)
Surface Contamination test: < 185 Bq (5 nCi) (6)

The undersigned, authorized officer of QSA Global, Inc. certifies that this source complies with the requirements of ISO2919 and that all of the information given in this certificate is true and correct.

Quality Assurance

Munk = 15 Aug 12

(1) Date Format yyyy-mm-dd.

(2) Confidence level of 99.7%.

(3) The apparent activity is determined by applying a conversion factor (0.110 mGy m².⁻¹GBq⁻¹) to the measured gamma radiation output of the sealed source determined with a calibrated instrument. The instrument is traceable to the National Institute of Standards & Technology. (NIST)

(4) The apparent activity is the Iridium-192 activity; other radionuclides not detectable.

(5) Leakage test method according to ISO9978 method Liquid nitrogen bubble test (6.2.4).

(6) Surface contamination test according to ISO9978 method Wet wipe test (5.3.1).

(7) The quality assurance system of QSA Global Inc. is certified by Lloyd's Register Quality Assurance (LRQA) according to ISO 9001 and ISO 13485.



QSA GLOBAL

40 North Avenue Burlington, MA 01803 1(781) 272-2000



Source Return Document (USA) for Container type ADIC (Part No. 081.040-xx)

DO NOT REMOVE THIS DOCUMENT FROM THE TRANSPORT DRUM

Hospital name:	SIBLEY MEMORIAL HOSPITAL	Address:	5255 LOUGHBORO RD. NEW
City:	WASHINGTON DC	State/Country:	DE 20016 USA
Container serial no. (on top of lid):	2303CG	Afterloader serial no.:	31511
Source serial no. (on source tail):	D36E1653		
Source type:	<input checked="" type="checkbox"/> mHDR V3 / V2 / Digital (0.9mm) <input type="checkbox"/> mHDR Classic (1.1mm) <input type="checkbox"/> mPDR V3 / V2 / Digital (0.9mm) <input type="checkbox"/> mPDR Classic (1.1mm)		

No.	Item (see Source Handling Procedure, part no. 090.387)	Nucletron Certified Source Handler	
1.	Are you Nucletron-certified for source exchanges?	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
2.	Dose rate above the source channel (without lid) should be lower than on the side of the container	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
3.	Source is locked (lever down) and source cable is undamaged	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
4.	A: Length of source cable extending from unload tube (Source is in channel 2 of the container)	$\leq 480\text{mm}$ (0.9 mm source)	$\leq 460\text{mm}$ (1.1 mm source)
		477	
	B: Source is in channel 1: push source to bottom of channel. Lock the source cable again (locking lever down)	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
5.	Source information label on container matches actual source	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
6.	Container lid is closed: both nuts fastened and lid locked	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
		$\leq 20\text{ mrem/h}$ or $\leq 200\text{ }\mu\text{Sv/h}$	
7.	Surface dose rate of the hot spot on top of the transport drum	1.6	
8.	Surface dose rate of the hot spot on the side of the transport drum	10.0	
9.	Dose rate for the hot spot at 1 meter from the transport drum	0.3	
10.	Radioactive contamination on transport drum	$\leq 0.4\text{Bq/cm}^2$	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Remark
11.	Measurements within specification?	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
12.	Transport Index (TI)	$0.1 \leq \text{TI} \leq 0.6$	0.3
13.	Source activity	150.82	GBq
14.	Labels are correctly placed and filled in (II-Yellow and address labels)	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
15.	Shipping documents are completed and signed	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
16.	Are there quality problems to be reported about: source, container(s), packaging, labeling, error code 227/228, etc.? If Yes, fill in the remark and attach the RMA label on the container and transport drum, see step 17.	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark
17.	RMA label attached?	<input checked="" type="checkbox"/> No	<input type="checkbox"/> Yes
18.	Lid screwed tightly on to transport drum and red/white tamper inserted; source return document enclosed (tick 'Pass' if done immediately after completing this form)	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Remark

Nucletron Certified Source Handler:	Signature:	Date (YYYY-MM-DD):
Rick Carlini		2012-AUG-27


Include the white original in the transport drum. For shipment authorization, fax or e-mail a copy to Nucletron: Fax no. (1)-410-312-4196 / e-mail: us_weeklyservicepackage@us.nucletron.com

Remarks: Place a RMA label on the container and on the transport drum.
--

Shipment Authorization for return shipment outside US & Canada <small>(to be filled in by authorized Nucletron employee)</small>	
Name:	<input type="checkbox"/> Approved
Signature:	<input type="checkbox"/> NO: DO NOT SHIP and call Nucletron (1)-800-234-2249
Date:	

Fourth Quarter Sealed Source Inventory

On November 5, 2012 inventory of the sealed sources in the new Brachy room was done by the undersigned. All sealed sources were transferred from the old department in a locked safe with exception of the HDR source which was stored inside the nucletron microSelectron mHDR unit. The HDR unit had a 384 GBq (10.4 Ci) Ir-192 source, S/N D36E-2572, installed on August 27, 2012. The previous source, S/N D36E-1653 was shipped out September 24, 2012. All brachytherapy seeds were noted and accounted for. A total of 561 seeds were inventoried of which 520 were I-125 seeds and 41 were Cs-131 seeds. All seeds were kept in lead containers in the safe below the L-block. These seeds are kept under lock and key. The door to the brachy room is always kept locked. An area survey of the countertop, safe, floor, and storage room across for the brachy room revealed all areas to be < 50 cpm using the Ludlum 14 C survey meter, S/N 206513. Background measured 50 cpm.

 11-12-12
Jordie Keck, M.S., DABR