Subject: FW: Response to Question from NRC

From: Spiezio, Melissa [mailto:MAA1@phin.org]
Sent: Wednesday, December 12, 2012 11:28 PM
To: Tripp, Lester
Cc: Keenan, James; Spiezio, Melissa
Subject: Response to Question from NRC

Licensee: Southwestern Vermont Medical Center

License No: 44-11345-02

Docket No: 03008027

Mail Control No: 578074

To: Lester Tripp

Date: December 12, 2012

This correspondence is in reference to our application dated July 23, 2012, requesting the renewal of U.S.NRC License No. 44-11345-02.

We confirm that we will perform a prospective evaluation demonstrating that the unmonitored individuals are not likely to receive, in any one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20, or we will provide dosimetry that meets the requirements listed under “Criteria” in NUREG-1556, Vol. 9, Rev 2 “Consolidated Guidance About Material Licenses: Program Specific Guidance About Medical Use Licensees,” dated October 2008. . This statement is true for licensees who use any of the major dosimetry services, e.g. Landauer and Siemans, Mirion, among others.

Please feel free to contact me with any further questions.

Thank you,

Melissa Spiezio

DISCLAIMER: The information contained in this electronic message is legally privileged and confidential under applicable law and is intended for a particular addressee(s). If it is not clear that you are the intended recipient, you are hereby notified that you have received this transmittal

in error; any review, copying or distribution or dissemination is strictly prohibited. If you suspect that you have received this transmittal in error, please notify Southwestern VT Health Care Corporation immediately by return email reply to (helpdesk@phin.org), and immediately delete the transmittal and any attachments without making any copy or distribution. Thank you.