



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
2100 RENAISSANCE BOULEVARD, SUITE 100
KING OF PRUSSIA, PENNSYLVANIA 19406-2713

December 20, 2012

Docket No. 03038594
Control No. 579495

License No. 20-35022-01

Randall S. Kaminsky
Sr. Vice President
CSMI, LLC
60 Island Street
Lawrence, MA 01840-1835

**SUBJECT: CSMI, LLC, REQUEST FOR ADDITIONAL INFORMATION CONCERNING
APPLICATION FOR NEW LICENSE, CONTROL NO. 579495**

Dear Mr. Kaminsky:

This is in reference to your letter dated December 3, 2012 (ADAMS Accession Number ML12340A385) applying for a Nuclear Regulatory Commission license. In order to continue our review, we need the following additional information:

1. Regarding Item 5 of your application, identify all the sealed sources you wish to service by radioisotope, manufacturer (or distributor), model number and the maximum amount of licensed material in the source. Confirm that each sealed source, device, and source/device combination is registered as an approved sealed source or device by NRC or an Agreement State. Confirm that the activity per source and maximum activity per device will not exceed the maximum activity listed on the approved certificate of registration issued by NRC or by an Agreement State. Refer to Section 8.5.1 of NUREG-1556, Vol. 18, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Serviced Provider Licenses" (NUREG-1556, Vol. 18).
2. Regarding Item 6 of your application, provide specific information regarding the purpose for which the licensed material will be used. Refer to Section 8.6 of NUREG-1556, Vol. 18.
3. Verify that the licensed material will be serviced at temporary job sites. Please provide the locations of the temporary job sites.
4. Regarding Item 7 of your application, please describe the RSO's SAIC Field Service Representative radiation safety training, including the location, dates, and duration of the training. Also, provide a more complete organizational chart showing the RSO's independence and direct communication with responsible management officials and demonstrating day-to-day oversight of the radiation safety activities. This information should not include personally identifiable information (PII), which is social security number, birth date, home phone number, etc. Please refer to Section 8.7.1 of NUREG-1556, Vol. 18.

5. Please verify that the authorized users have received and/or will receive training regarding each specific device to be serviced. Please provide a description of the training and the dates of training. This information should not include PII. Please refer to Section 8.7.2 of NUREG-1556, Vol. 18.
6. List any other classification of personnel who will require radiation safety training and describe the radiation safety training program, including topics covered, groups of workers, assessment of training, qualifications of instructors, and the method and frequency of training, that will be provided to individuals working in or frequenting restricted areas. Please refer to Sections 8.7.3 and 8.8 of NUREG-1556, Vol. 18.
7. Regarding leak testing and analysis, please state that you will commit to Appendix O of NUREG-1556, Vol. 18.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 579495. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5399.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

The NRC's Safety Culture Policy Statement became effective in June 2011. While a policy statement and not a regulation, it sets forth the agency's *expectations* for individuals and organizations to establish and maintain a positive safety culture. You can access the policy statement and supporting material that may benefit your organization on NRC's safety culture Web site at <http://www.nrc.gov/about-nrc/regulatory/enforcement/safety-culture.html>. We strongly encourage you to review this material and adapt it to your particular needs in order to develop and maintain a positive safety culture as you engage in NRC-regulated activities.

Sincerely,

Original signed by John Nicholson for

Stephen Hammann
Senior Health Physicist
Decommissioning Branch
Division of Nuclear Materials Safety

cc:

R. Kaminsky

3

Roberto Bahday, Radiation Safety Officer

DOCUMENT NAME: G:\WordDocs\Current\Lic Def Letter\L20-35022-01.579495.doc

SUNSI Review Complete: LKauffman

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	DNMS/RI		
NAME	LKauffman/LKauffman		SHammann/JM for			
DATE	12/ 20 /2012		12/ 20 /2012			

OFFICIAL RECORD COPY