

U.S. NUCLEAR REGULATORY COMMISSION

Date: 1/2/2013

TELEPHONE CONVERSATION RECORD

| | | | | | |
|-------------------------|--------|-----------------------|--------------|----------------------|----------|
| Mail Control | 579158 | License No(s). | 07-28762-01 | Docket No(s). | 03032792 |
| or Report No(s). | 579159 | | 07-28762-02G | | 03032988 |

Name of Licensee: Agilent Technologies, Inc.

Name of Participant(s): David Bennett, RSO
Dennis Lawyer, R1, DNMS

Telephone No. 302-633-8262

Subject: RAI, Proprietary Information and Sealed Source Information
(NOTE: This will be used as the Documents Title in ADAMS)

Mr. Bennett, In order for your letter dated October 1, 2012 to be considered proprietary, you need to submit an affidavit per 10 CFR 2.390. I do not need anything proprietary and would suggest you submit leak test procedures without proprietary information or state you will follow Appendix R of NUREG-1556, Volume 7. For the temporary job locations, I needed you to state that you have security procedures in place to prevent loss and theft and the method to prevent contamination spread. Since you are only using foil sources, performing the leak test within the prescribed frequency on the sealed source and device (SS&D) registry would be sufficient. On the general distribution listing, many of the sources being listed were not on the SS&D registry. Also two of these are only to be distributed to specific licensees and not general licensees.

Action Required: Mr. Bennett will submit corrections and the additional information needed.

Document Availability: **Publicly Available** **Non-Publicly Available**

Non-Sensitive **Non-Sensitive Copyright** **Sensitive- Proprietary**

Sensitive – Privacy Act/ PII **Sensitive – Internal** **Sensitive – Security-Related**

Immediate Release **Normal Release Date: 01/10/2013** **Delayed Release Date:**

SUNSI Review Completed By: Dennis Lawyer / RA /