

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: St. Louis Cardiology Consultants 11133 Dunn Road Suite 2346 St. Louis, MO 63136 REPORT NUMBER(S) 13-01	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352	
3. DOCKET NUMBER(S) 030-36070	4. LICENSE NUMBER(S) 24-32413-01	5. DATE(S) OF INSPECTION January 14, 2013

LICENSEE:
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

1. Based on the inspection findings, no violations were identified.

2. Previous violation(s) closed.



3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
 (Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		1/14/13
BRANCH CHIEF	Tamara E. Bloomer		1/27/13

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: St. Louis Cardiology Consultants 11133 Dunn Road Suite 2346 St. Louis, MO 63136 REPORT NUMBER(S) 13-01	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
---	---

3. DOCKET NUMBER(S) 030-36070	4. LICENSE NUMBER(S) 24-32413-01	5. DATE(S) OF INSPECTION January 14, 2013
--------------------------------------	---	--

6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01-03.07
--	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT S. Narayanan, M.D., RSO	4. TELEPHONE NUMBER (314) 653-5125
---------------------------------	----------------------	--	---

Main Office Inspection Next Inspection Date: 01/12/2018

Field Office Inspection 261 Dunn Road, Florissant, MO

Temporary Job Site Inspection

PROGRAM SCOPE

The licensee was a cardiac clinic authorized by the license to use as needed, any byproduct material permitted by 10 CFR 35.100 and 35.200 at the location specified on the license.

The nuclear medicine department was staffed with one nuclear medicine technologist (NMT). The licensee performed only cardiac studies and averaged from none to 3 patient studies scheduled on Monday, Tuesday, Thursday, and Friday each week. The nuclear medicine department received rest and stress unit doses from a local St. Louis, MO nuclear pharmacy as ordered. All waste was either held for decay-in-storage (DIS) or returned to the nuclear pharmacy as limited quantity shipments.

Performance Observations

The licensee's NMT demonstrated/discussed: (1) survey meter use and calibrations; (2) package check-in procedures; (3) unit dosage prep and safe use; (4) wipe test counting; (5) waste handling; (6) sealed source inventories and leak tests; (7) routine security of licensed material; (8) dose calibrator tests; (9) radiation safety program audits; (10) any contamination events (none); (11) HAZMAT refresher training; and (12) dosimetry records indicated occupational dose for the NMT less than 10% of the annual limits.

The inspector performed independent and confirmatory radiation measurements which indicated results consistent with licensee survey records and postings.