



Copley Hospital

528 Washington Highway
Morrisville, VT 05661
(802) 888-4231
www.copleyvt.org

Br. 1

REC RG 1 01 31 13 AM 07 18

US NRC Region I DNMS
2100 Renaissance Blvd.
King of Prussia, PA 19406

January 24, 2013

Dear Sir or Madam,

03017125

I am requesting an amendment of our current NRC Byproducts Materials license, #44-19196-01, for the addition of Gadolinium-153. We have purchased a new SPECT/CT gamma camera that requires the use of a Gadolinium-153 line source, 10mCi in activity, for routine calibration purposes. Gd-153 is not listed for use on our current license.

The Gd-153 line source is for the Automatic Quality Control (AQC) device used on this equipment which has received 510(k) clearance by the FDA under registration #K041166. The AQC device was approved by the Illinois Emergency Management Agency, Division of Nuclear Safety, under device registry #IL-605-D-106-S, which is acknowledged by the NRC.

If you need further information please do not hesitate to contact me.

Thank you for your assistance in this matter.

Sincerely,

Melvyn Patashnick, Chief Executive Officer

Expert Care with a Personal Touch

579929
NMSS/RGN1 MATERIALS-002

This is to acknowledge the receipt of your letter application dated

11/24/13, and to inform you that the initial processing which includes an administrative review has been performed.

Amendment (44-19196-01) There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 579929.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.