

U.S. NUCLEAR REGULATORY COMMISSION

Date: 02/13/2013

TELEPHONE CONVERSATION RECORD

Mail Control or Report No(s): 579828 License No(s): 06-14001-01 Docket No(s): 03003850

Name of Licensee: Ahlstrom Nonwovens LLC
Name of Participant(s): Timothy Daly, RSO
Scott Wilson, RI Health Physicist
Telephone No. 860-654-8560
Subject: RAI License Renewal App. dtd 01/02/2013

Licensee was contacted to request additional information regarding the license renewal application. Specifically, the following questions were asked:

1. How many gauges are possessed? 3 Model TG-4 fixed gauges with a maximum of 1.0 Ci (Kr-85) each.
2. What are the manufacturer and model numbers? ABB, Inc. Model TG-4. The model TG-1 gauges were returned to the manufacturer and replaced with TG-4 models.
3. Section 10 (Maintenance) of Appendix B of the application does not include an answer to the question regarding Non-Routine Maintenance Operations. RSO stated that the licensee does not perform non-routine maintenance. Non-Routine maintenance is performed by the manufacturer.
4. Maximum possession limit is 2.5 Ci on the current license. Licensee requested an increase in the possession limit to 4 Ci to allow possession of three 1 Ci sources, and an additional source possession if a replacement is received.

The licensee was asked to revise the application and submit the revised document to RI via e-mail, fax, or mail. Licensee agreed to submit the revised application within 7 days.

Action Required: Save this record in ADAMS referencing the Docket, Mail Control and license numbers above. Licensee to submit revised license renewal application within 7 days.

Document Availability:

 Publicly Available Non-Publicly Available Non-Sensitive
 Sensitive – Privacy Act/ PII
 Immediate Release Non-Sensitive Copyright
 Sensitive – Internal
Normal Release Date: 02/22/2013 Sensitive- Proprietary
 Sensitive – Security-Related
Delayed Release Date:SUNSI Review Completed By: SWilson / RA /