

UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352



NOV 16 2012

David Schmitt, R.Ph.
Radiation Safety Officer
Lakeview Diagnostic, L.L.C.
36211 Jefferson Avenue
Harrison Township, MI 48045

Dear Mr. Schmitt:

This refers to your letter dated October 29, 2012, which requests closing the Lakeview Diagnostic facility in Port Huron under license no. 21-32444-01.

We cannot authorize licensees to release the locations of use from licenses for unrestricted use (even by other members of their staff) until we have received and reviewed a copy of the results of decommissioning and close-out surveys for the facilities.

The final status survey must include a complete historical review of all actual licensed materials used, including sealed and unsealed sources, spills, and contamination.

If sealed sources were transferred as part of the close-out of this license, please provide a copy of the final leak test for each sealed source, prepared in accordance with 10 CFR 35.2067; a copy of an acknowledgment of receipt from the licensed entity who took possession of each source; and if the recipient/transferee is an Agreement State license, please include a current copy of its license that clearly shows it is licensed to receive your sources.

Your complete historical review should specify when and where all licensed materials, especially which materials in Subitem No. 6A., were possessed, used and/or stored, when the last use was for licensed materials and how, when and by whom were the materials disposed of (shipped off site, decayed -in-storage, sanitary sewer disposal, etc.) or transferred.

The final status survey should consist of exposure rate measurements to show that all sources of radioactive material have been removed, and contamination checks of areas where radioactive materials were used or stored. Average radiation levels associated with surface contamination and removable contamination should not exceed those specified in the decontamination guide being sent to you under separate cover.

Please also refer to section 15.5.3 in NUREG 1757, Vol. 1, Rev. 1, available on our website, for additional assistance.

Please submit the following information with your close-out survey:

- a. Diagrams of each facility with exposure rate survey and wipe test results keyed to specific locations, as appropriate.

D. Schmitt

- b. The name of the person performing the survey.**
- c. The date the survey was performed.**
- d. The instrument(s) used for exposure rate measurements and for analysis of the wipes.**
- e. Background readings and each instruments' efficiency or correction factor.**
- f. The date(s) that the survey instruments were last calibrated.**
- g. The action levels for both exposure rate measurements and wipe tests. Include the identity of areas exceeding these levels, corrective actions taken and results of corrective actions taken.**

Final copies of radioactive waste stream disposals from decay-in-storage should also be submitted in support of your request.

Please complete a Form NRC 314 also and include it with your response so it is clear that your intentions are to terminate this license. It was not clear from your letter dated October 29, 2012, that termination of this license was what you intended.

These references may also assist you: 10 CFR 30.41; 10 CFR 30.51; NUREG 1556 Vol. 9, Rev. 2, section 11, "Termination of Activities," and Appendix H, "NRC Form 314."

Please always include the telephone number and fax number of at least one person who serves as a point of contact for all future licensing requests.

Please submit your response in writing, within 30 days of the date of this letter, or contact me to make alternative response arrangements. Please address your response to my attention at the address above as "additional information to control number 579340."

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,



Colleen Carol Casey
Materials Licensing Branch

License No. 21-32444-01MD
Docket No. 030-36222