



MILFORD HOSPITAL

Milford Hospital  
300 Seaside Ave.  
Milford CT. 06460


License No.: 06-13611-01  
Docket No: 030-01298  
Control No: 579550

Here are the responses to your questions for additional data.

1. We do not use any PET Radiopharmaceuticals at this institution.
2. Please revise my initial response for section 10.4 to read as follows:  
Safe use of Unsealed Licensed Material: We have developed and will implement and maintain written procedures for safe use of unsealed byproduct material that meet the requirements of 10.CFR 20.1101 and 10CFR 20.1301.
3. Please add to section 9.2 the following instruments currently used at our Facility. Please note that the models and serial numbers may change over time as instruments are added or deleted from use.  
Captus 3000 Thyroid probe S/N 200358/Well counter (separate chamber) S/N 115865  
Geiger counters: Ludlum 14C S/N 95924; Bicon MicroAnalyst S/N B013R;  
Ludlum 14-C-2 S/N 129033; and Victoreen 190 S/N 1356  
Dose Calibrator: CRC-25R S/N 251231

This response was drafted by the Radiation Safety Officer James Ingarra.

Joseph Pelaccia, President

 Date: 2 18 13