

Sara A.B. Forster
MATERIALS LICENSING BRANCH

TELECON & FAX TRANSMITTAL



TO: file
COMPANY: N/A

NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD
LISLE, ILLINOIS 60532-4351
(630) 829-9892 FAX: (630) 515-1078

PAGES: 2 TEL.: (269) 373-7407

FAX #: N/A
EMAIL: N/A

CONVERSATION RECORD

CONVERSATION RECORD		TIME	DATE
		3:45 pm	January 7 & February 6, 2013
NAME OF PERSON(S) CONTACTED	TELEPHONE NO.	ORGANIZATION	
Paul Jursinic, Ph.D., RSO	(269) 373-7407	West Michigan Cancer Center	
REPRESENTED PERSON or PERSONS		ORGANIZATION	
Paul Jursinic, Ph.D., Radiation Safety Officer		West Michigan Cancer Center	
SUBJECT			
License No.: 21-32501-01		Control No.: 579406	

SUMMARY

We have reviewed your November 6, 2012, license amendment request, to add Jeffrey Daniel Radawski as an Authorized User (AU) for 10 CFR 35.600 (HDR, Ir-192) materials, and find that we are unable to continue this action until we have received information regarding the following:

1. The NRC Form 313A (AUS) indicated only a general date range for Dr. Radawski's classroom and lab training. It included neither specific dates on which training was received nor residency program confirmation all topics were covered during the program. **Resubmit NRC Form 313A (AUS). Include specific training dates, and/or residency program documentation confirming that 10 CFR 35.690(b)(1)(i) required topics required were covered over a minimum of 200 hours, as part of the standard residency program. Please include a copy of Dr. Radawski's residency program completion certificate, in support of your request.**
2. The work experience outlined on the submitted NRC Form 313A (AUS) included 10 CFR 35.490 experience, as well as a references to 10 CFR 35.690 (HDR, teletherapy & GSR) experience. Please note that the 10 CFR 35.490 and the 10 CFR 35.690 (teletherapy & GSR) experience were not considered in this review, as the only authorized medical use listed on your licenses is under 10 CFR 35.600 (HDR). Regarding the 10 CFR 35.690 (HDR) request, the NRC Form 313A (AUS) listed only the general date range over which specific topics were covered, and did not include Dr. Radawski's total hours of experience. **Please include specific work experience dates and total hours of experience confirming that Dr. Radawski has completed the 10 CFR 35.690(b)(1)(ii) 500 hours minimum work experience under supervision of an approved AU, in the resubmitted NRC Form 313A (AUS). If work experience was completed Christopher E. Pelloski, M.D., provide a copy of the Ohio State University Department of Radiation Oncology license, license 02110250037, and a confirmatory letter from that licensee's RSO or RSC chair indicating that Dr. Pelloski was approved as the equivalent of a 35.690 (HDR) AU under that license. Documentation provided should clearly indicate the device manufacturer and model number on which any procedures were performed.**

- 3. No supervising individual was listed for specific remote afterloader training listed on the submitted NRC Form 313A (AUS). Some training dates were also missing from that form. **Please include the dates on which 10 CFR 35.690(c) HDR device operation and safety procedures training was completed, in your resubmitted NRC Form 313A (AUS). List several specific dates on which HDR clinical use training was completed. Include the name(s) of any supervising physician AUs, but do not include any patient names or other privacy protected information with your response. If the device on which training was obtained is different than Licensee's Varian GammaMed Plus, please confirm that Dr. Radawski has been or will be trained on the general device operation and emergency procedures for the licensee's GammaMed device.**

We have requested that you submit referenced items to add AU Jeffrey Daniel Radawski, M.D.

- Classroom and lab training, including residency program details and certificate
- NRC Form 313A (AUS), including supervised work experience under an AU
- HDR training details including device-specific training confirmation

– via facsimile, to (630) 515-1078. Please reference Control No. 579406, as listed at the top of this memo. We had requested to hear from the licensee on or before February 6, 2013.

However, in our February 6, 2013 phone conversation, the licensee contact person indicated that the requested training documentation for Dr. Radawski had been requested from the training institution but had not yet been received. The licensee stated that it is best to table the request until a response is available. The action will have been voided without prejudice, accordingly.

For future reference, please always include the name, phone number and fax number of at least one person whom we may contact for additional information when reviewing your licensing correspondence and requests.

Please submit the requested information within **30** days of this record. **Include reference control number 579406, and a signed, dated cover letter with your response. Please FAX your response to my attention at (630) 515-1078.** You may also scan your response and send to me via email, as a pdf file.

Please direct any questions you have to me at **(630) 829-9892** or **sara.forster@nrc.gov**.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Sara A.B. Forster

Sara A.B. Forster 02/11/2013