



Designer and Manufacturer
of
Scientific and Industrial
Instruments

CERTIFICATE OF CALIBRATION

LUDLUM MEASUREMENTS, INC.
POST OFFICE BOX 810 PH. 325-235-5494
501 OAK STREET FAX NO. 325-235-4672
SWEETWATER, TEXAS 79556, U.S.A.

CUSTOMER INSTITUTE FOR PHARMA DISCOVERY ORDER NO. 20190662/371799
Mfg. Ludlum Measurements, Inc. Model 3 Serial No. 153923
Mfg. Ludlum Measurements, Inc. Model 44-9 Serial No. PR157495
Cal. Date 20-Dec-11 Cal Due Date 20-Dec-12 Cal. Interval 1 Year Meterface 202-608

Check mark Applies to applicable instr. and/or detector IAW mfg. spec. T. 72 °F RH 28 % Alt 797.8 mm Hg

New Instrument Instrument Received Within Toler. +/-10% 10-20% Out of Tol. Requiring Repair Other-See comments

Mechanical ck. Meter Zeroed Background Subtract Input Sens. Linearity

F/S Resp. ck. Reset ck. Window Operation Geotropism

Audio ck. Alarm Setting ck. Batt. ck. (Min. Volt) 2.2 VDC

Calibrated in accordance with LMI SOP 14.8 rev 12/05/89. Calibrated in accordance with LMI SOP 14.9 rev 02/07/97.

Instrument Volt Set 900 V Input Sens. 34 mV Det. Oper. 900 V at 34 mV Threshold Dial Ratio = _____ mV

HV Readout (2 points) Ref./Inst. _____ / _____ V Ref./Inst. _____ / _____ V

COMMENTS:

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING*
X 100	150 mR/hr	1.5	1.5
X 100	50 mR/hr	0.5	0.5
X 10	15 mR/hr	1.5	1.5
X 10	5 mR/hr	0.5	0.5
X 1	1.5 mR/hr = <u>4930 cpm</u>	1.5	1.5
X 1	1.0 mR/hr	1.0	1.0
X 0.1	<u>493</u> cpm	1.5	1.5
X 0.1	<u>164</u> cpm	0.5	0.5

*Uncertainty within ± 10% C.F. within ± 20%

X 0.1 Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*
Digital Readout			Log Scale		

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements of ANSI/NCSS Z540-1-1994 and ANSI N323-1978

State of Texas Calibration License No. LO-1963

Reference Instruments and/or Sources: 73410 1131 781 059 280 60646 70897 Ra-226 S/N Y982
Cs-137 Gamma S/N 1162 G112 M565 5105 T1008 T879 E552 E551 720 734 1616 Neutron Am-241 Be S/N T-304
 Alpha S/N _____ Beta S/N _____ Other _____
 m 500 S/N 81084 Oscilloscope S/N _____ Multimeter S/N 69101832

Calibrated By: Donald E. Brenner Date 20-DEC-11

Reviewed By: [Signature] Date 20 Dec 11

This certificate shall not be reproduced except in full, without the written approval of Ludlum Measurements, Inc.
FORM C22A 10/24/2011 Page 1 of 1

AC Inst. Only Passed Dielectric (Hi-Pot) and Continuity Test Failed: _____



Designer and Manufacturer
of
Scientific and Industrial
Instruments

CERTIFICATE OF CALIBRATION

LUDLUM MEASUREMENTS, INC.
POST OFFICE BOX 810 PH. 325-235-5494
501 OAK STREET FAX NO. 325-235-4672
SWEETWATER, TEXAS 79556, U.S.A.

ORDER NO. 20192163/372715

CUSTOMER INSTITUTE OF PHARMA DISCOVERY

Mfg. Ludlum Measurements, Inc. Model 3

Serial No. 155003

Mfg. Ludlum Measurements, Inc. Model 44-9

Serial No. PR 157465

Cal. Date 20-Jan-12 Cal Due Date 20-Jan-13

Cal. Interval 1 Year Meterface 202-608

T. 75 °F RH 20 % Alt 692.8 mm Hg

Check mark applies to applicable instr. and/or detector IAW mfg. spec.

New Instrument Instrument Received Within Toler. +/-10% 10-20% Out of Tol. Requiring Repair Other-See comments

Mechanical ck. Meter Zeroed Background Subtract Input Sens. Linearity

F/S Resp. ck. Reset ck. Window Operation Geotropism

Audio ck. Alarm Setting ck. Batt. ck. (Min. Volt) 2.2 VDC

Calibrated in accordance with LMI SOP 14.8 rev 12/05/89. Calibrated in accordance with LMI SOP 14.9 rev 02/07/97.

Instrument Volt Set 900 V Input Sens. 34 mV Def. Oper. 900 V at 34 mV Threshold Dial Ratio = mV

HV Readout (2 points) Ref./Inst. / V Ref./Inst. / V

COMMENTS:

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING*
X 100	150 mR/hr	1.5	1.5
X 100	50 mR/hr	0.5	0.5
X 10	15 mR/hr	1.5	1.5
X 10	5 mR/hr	0.5	0.5
X 1	1.5 mR/hr = 4680 cpm	1.5	1.5
X 1	1.0 mR/hr	1.0	1.0
X 0.1	468 cpm	1.5	1.5
X 0.1	156 cpm	0.5	0.5

*Uncertainty within ± 10% C.F. within ± 20%

X 0.1 Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*
Digital Readout			Log Scale		

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other international Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements of ANSI/NCSL Z540-1-1994 and ANSI N323-1978. State of Texas Calibration License No. LO-1963

Reference Instruments and/or Sources: 73410 1131 781 059 280 60646 70897 Ra-226 S/N Y982
Cs-137 Gamma S/N 1162 G112 M565 5105 T1008 T879 E552 E551 720 734 1616 Neutron Am-241 Be S/N T-304
 Alpha S/N Beta S/N Other m 500 S/N 189506 Oscilloscope S/N Multimeter S/N 94000441

Calibrated By: William Tinsley

Date 20-January-2012

Reviewed By: [Signature]

Date 20 Jan 12

AC Inst. Only Passed Dielectric (Hi-Pot) and Continuity Test Failed:



Designer and Manufacturer
of
Scientific and Industrial
Instruments

CERTIFICATE OF CALIBRATION

LUDLUM MEASUREMENTS, INC.
POST OFFICE BOX 810 PH. 325-235-5494
501 OAK STREET FAX NO. 325-235-4672
SWEETWATER, TEXAS 79556, U.S.A.

CUSTOMER INSTITUTE FOR BIOANALYTICS

ORDER NO. 20189779/371291

Mfg. Ludlum Measurements, Inc. Model 3

Serial No. 154510

Mfg. Ludlum Measurements, Inc. Model 44-9

Serial No. PR157474

Cal. Date 6-Dec-11 Cal Due Date 6-Dec-12 Cal. Interval 1 Year Meterface 202-608

Check mark applies to applicable instr. and/or detector IAW mfg. spec. T. 72 °F RH 20 % Alt 706.8 mm Hg

- New Instrument Instrument Received Within Toler. +/-10% 10-20% Out of Tol. Requiring Repair Other-See comments
- Mechanical ck. Meter Zeroed Background Subtract Input Sens. Linearity
- F/S Resp. ck. Reset ck. Window Operation Geotropism
- Audio ck. Alarm Setting ck. Batt. ck. (Min. Volt) 2.2 VDC
- Calibrated in accordance with LMI SOP 14.8 rev 12/05/89. Calibrated in accordance with LMI SOP 14.9 rev 02/07/97.

Instrument Volt Set 900 V Input Sens. 34 mV Det. Oper. 900 V at 34 mV Threshold Dial Ratio = _____ mV

HV Readout (2 points) Ref./Inst. _____ / _____ V Ref./Inst. _____ / _____ V

COMMENTS:

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING*
X 100	150 mR/hr	1.5	1.5
X 100	50 mR/hr	0.5	0.5
X 10	15 mR/hr	1.5	1.5
X 10	5 mR/hr	0.5	0.5
X 1	1.5 mR/hr = 5000 cpm	1.5	1.5
X 1	1.0 mR/hr	1.0	1.0
X 0.1	500 cpm	1.5	1.5
X 0.1	167 cpm	0.5	0.5

*Uncertainty within ± 10% C.F. within ± 20%

X 0.1 Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*	Log Scale	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*
_____	_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____	_____

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements of ANSI/NCSL Z540-1-1994 and ANSI N323-1978

State of Texas Calibration License No. LO-1963

- Reference Instruments and/or Sources: 73410 1131 781 059 280 60646 70897 Ra-226 S/N Y982
- Cs-137 Gamma S/N 1162 G112 M565 5105 T1008 T879 E552 E551 720 734 1616 Neutron Am-241 Be S/N T-304
- Alpha S/N _____ Beta S/N _____ Other _____
- m 500 S/N 81084 Oscilloscope S/N _____ Multimeter S/N 69101832

Calibrated By: Donald E. Brennan Date 6-DEC-11

Reviewed By: [Signature] Date 6 Dec 11

AC Inst. Passed Dielectric (Hi-Pot) and Continuity Test
Only Failed: _____