



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
2100 RENAISSANCE BOULEVARD, SUITE 100
KING OF PRUSSIA, PENNSYLVANIA 19406-2713

March 4, 2013

Docket No. 03038609
Control No. 579878

License No. 06-35037-01

John Farrelly, DVM
Veterinary Oncology and Hematology Center, LLC
178 Connecticut Avenue
Norwalk, CT 06854

SUBJECT: VETERINARY ONCOLOGY AND HEMATOLOGY CENTER, LLC, REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION FOR AMENDMENT TO LICENSE, CONTROL NO. 579878

Dear Dr. Farrelly, DVM:

This is in reference to your application dated January 21, 2013 and additional information you provided February 7, 2012 requesting an Nuclear Regulatory Commission License. In order to continue our review, we need the following additional information:

1. In your application you refer to an Emergency Plan (see Appendix 1 Below), however, the Appendix is missing.
2. In your application under training you refer to the attached lecture slides and handouts but they are not attached.
3. Please provide specific facility drawings and layouts with dimensions for your facilities where byproduct materials will be used or stored.
4. In your application you indicate audits are not applicable, however, annual audits are required by 10 CFR 20.1101(c). Please confirm you will conduct an annual audit.
5. Please provide additional details on the scope of activities of use for the ophthalmic applicator. What animals and applications will be used?

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits**, see our **toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 579878. If you have any technical questions

J. Farrelly, DVM

2

regarding this deficiency letter, please call me at (610) 337-5303.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Thomas K. Thompson

Thomas K. Thompson
Senior Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

DOCUMENT NAME: G:\WordDocs\Current\Lic Def Letter\L06-35037-01.579878.doc

SUNSI Review Complete: TThompson

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	DNMS/RI		
NAME	TThompson /TKT/					
DATE	3/4/2913					

OFFICIAL RECORD COPY