

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

| | | | |
|--|-------------------------------------|--|--|
| 1. LICENSEE/LOCATION INSPECTED: Jefferson Radiology, PC 85 Seymour Street, Suite 200 Hartford, Connecticut 06106 REPORT NUMBER(S) 2013-001 | | 2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 2100 Renaissance Blvd, Suite 100 King of Prussia, Pennsylvania 19406-2713 | |
| 3. DOCKET NUMBER(S) 030-31642 | 4. LICENSE NUMBER(S) 06-28502-01 | 5. DATE(S) OF INSPECTION March 26-27, 2013 | |

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☐ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s) and corrective action(s):

- ☒ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)

10 CFR 35.60(b) requires in part that a licensee calibrate instrumentation to measure the activity of unsealed byproduct material before it is administered to each patient or human research subject in accordance with nationally recognized standards or the manufacturer's instructions. Contrary to the above, on May 11, 2010, the licensee did not calibrate instrumentation to measure the activity of unsealed byproduct material before it is administered to each patient or human research subject in accordance with nationally recognized standards or the manufacturer's instructions. Specifically, the licensee installed a dose calibrator on May 11, 2010, but did not perform a test for geometric independence until March 27, 2013. Corrective actions included performing a geometry test on March 27, 2013, and retraining staff on dose calibrator requirements.

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

| Title | Printed Name | Signature | Date |
|---------------------------|--------------------------------|--------------------------------|---------|
| LICENSEE'S REPRESENTATIVE | Doreen Freeman | Doreen Freeman V.P. | 3/27/13 |
| NRC INSPECTOR | Janice Nguyen/Robert Gallagher | Janice Nguyen/Robert Gallagher | 3-27-13 |
| BRANCH CHIEF | James Dwyer | [Signature] | 4/23/13 |

