

**U.S. NUCLEAR REGULATORY COMMISSION**

Date:

5/3/2013

**TELEPHONE CONVERSATION RECORD**Mail Control  
or Report No(s).

580210

License No(s).

47-23065-01

03020231

Name of Licensee:

Thrasher Engineering, Inc.

Name of Participant(s):

Lee Gustafson, Architect, Thrasher Engineering  
Randolph C. Ragland, Jr., NRC Region I

Telephone No.

304-624-4108

Subject:

Thrasher Engineering Questions emailed to RidsRgn2MailCenter Resource

(NOTE: This will be used as the  
Documents Title in ADAMS)

Wade Loo, NRC Region II, forwarded an email that was sent to Region II from an Architect working for Thrasher Engineering (Thrasher), a Region I portable gauge licensee. The individual was requesting guidelines for designing a gauge storage room. He indicated that his company was moving and the new storage location will have room for 100 gauges. I informed him:

- 1) Thrasher will need a license amendment if they want to add additional gauges to their license;
- 2) Thrasher will need a license amendment if they plan to move the gauges to a new storage location;
- 3) The storage method must meet the security requirements of 10 CFR 30.34(i);
- 4) Thrasher can consult NUREG 1556, Vol. 1, for additional guidance (e.g., Appendix I);
- 5) The Thrasher Radiation Safety Officer, Andrew Kincell, can provide additional guidance; and
- 6) Thrasher can contact Randy Ragland (610-337-5083) or Scott Wilson (610-337-5136) with additional questions.
- 7) Due to a potential significant increase in possession limits, this action would likely prompt an inspection by NRC Region I.

Action Required: None

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

 Sensitive – Privacy Act/ PII Immediate Release

Non-Sensitive Copyright

Sensitive – Internal

Normal Release Date: 05/13/2013

Sensitive- Proprietary

Sensitive – Security-Related

Delayed Release Date:

SUNSI Review Completed By:

R. Ragland

/ RA /